

HEALTH MONITORING SYSTEM FOR AUTISM

Submitted in partial fulfilment of requirements for the award of
Bachelor of Engineering Degree in Electronics and Communication Engineering

by

SRIHARI V R (40130223)

SITHESHWAR B I (40130220)



**DEPARTMENT OF ELECTRONICS AND COMMUNICATION ENGINEERING
SCHOOL OF ELECTRICAL AND ELECTRONICS**

SATHYABAMA

**INSTITUTE OF SCIENCE AND TECHNOLOGY
(DEEMED TO BE UNIVERSITY)**

CATEGORY -1 UNIVERSITY BY UGC

**Accredited with Grade “A++” by NAAC I 12B Status by UGC I Approved by AICTE
JEPPIAAR NAGAR, RAJIV GANDHI SALAI, CHENNAI - 600 119**

APRIL – 2024



SATHYABAMA

INSTITUTE OF SCIENCE AND TECHNOLOGY

(DEEMED TO BE UNIVERSITY)

CATEGORY-I UNIVERSITY BY UGC

Accredited "A++" Grade by NAAC | 12B Status by UGC | Approved by AICTE

www.sathyabama.ac.in

DEPARTMENT OF ELECTRONICS AND COMMUNICATION ENGINEERING

BONAFIDE CERTIFICATE

This is to certify that this Project Report is the bonafide work of SRIHARI V R (40130223), SITHESHWAR B I (40130220), who have done the Project work as a team who carried out the project entitled "HEALTH MONITORING SYSTEM FOR AUTISM" under my supervision from November 2023 to April 2024.

Internal Guide

Dr. S. LALITHAKUMARI, M.E., Ph.D.

Head of the Department

Dr. T. RAVI, M.E., Ph.D.

Submitted for Viva voce Examination held on 18.04.2024

Internal Examiner

External Examiner

DECLARATION

I, **SITHESHWAR B I (40130220)** hereby declare that the Project Report entitled as **"HEALTH MONITORING SYSTEM FOR AUTISM"** done by me under the guidance of **Dr. S. LALITHAKUMARI, M.E., Ph.D., Department of Electronics and Communication Engineering**, at Sathyabama Institute of Science and Technology is submitted in partial fulfillment of the requirements for the award of Bachelor of Engineering degree in Electronics and Communication Engineering.

DATE: 18.04.2024

PLACE: CHENNAI

SIGNATURE OF THE CANDIDATES

1. V.B. Sri Hari

2. B.I. Sitheswar

DECLARATION

I, **SRIHARI V R (40130223)** hereby declare that the Project Report entitled as **"HEALTH MONITORING SYSTEM FOR AUTISM"** done by me under the guidance of **Dr. S. LALITHAKUMARI, M.E., Ph.D., Department of Electronics and Communication Engineering**, at Sathyabama Institute of Science and Technology is submitted in partial fulfillment of the requirements for the award of Bachelor of Engineering degree in Electronics and Communication Engineering.

DATE: 18.04.2024

PLACE: CHENNAI

SIGNATURE OF THE CANDIDATES

1. V.R. Sri Hari

2. B.I. Sitheshwar

ACKNOWLEDGEMENT

I am pleased to acknowledge my sincere thanks to **The Board of Management of Sathyabama Institute of Science and Technology** for their kind encouragement in doing this project and for completing it successfully. I am grateful to them.

I convey my thanks to **Dr. N. M. NANDHITHA, M.E., Ph.D., Professor & Dean, School of Electrical and Electronics** and **Dr. T. RAVI, M.E., Ph.D., Professor & Head of the Department of Electronics and Communication Engineering** for providing me with the necessary support during the progressive reviews.

I would like to express my sincere and deep sense of gratitude to my Project Guide, **Dr. S. LALITHAKUMARI, M.E., Ph.D., Department of Electronics and Communication Engineering**, for her valuable guidance, suggestions and constant encouragement paved the way for the successful completion of my project work.

I wish to express my thanks to all Teaching and Non-teaching staff members of the Department of Electronics and Communication Engineering who were helpful in many ways for the completion of the project.

ABSTRACT

Health monitoring system plays an important role in telemedicine nowadays, and they replace manual monitoring. This project proposes a novel system that can overcome the difficulties encountered when a health care system is designed for autism-affected children and people, who find difficulty expressing themselves. This work aims to support people in expressing themselves and communicating. The system can continually monitor behavior patterns and vital indicators. It then looks for changes from baseline behavior that might indicate health problems. Bio-signals from the people are being collected through various sensors, such as an oximeter and heart beat sensor and these are processed by a suitable processor and transmitted to the dedicated app, which is designed for this particular application, and that is being suitably communicated to health care professionals. The entire circuitry is fabricated as a single module that can be worn by humans, and the data that is retrieved from the proposed model has been successfully monitored and is being used for future reference. This project combines a special sensor called MAX30102 with a smart platform called Blynk to make a system that watches over important health signs like body temperature, heart rate, and oxygen levels. The sensor is like a tiny doctor that measures these signs accurately. It talks to a small computer called a microcontroller, which collects and processes the data. The Blynk platform enables to transform health information through user's phone or computer in real-time. It can mimic the presence of a virtual doctor who can be vigilant on user's and can attend any sort of emergency needs.

TABLE OF CONTENTS

Chapter No.	TITLE	Page No.
	ABSTRACT	v
	LIST OF FIGURES	viii
1	INTRODUCTION	1
	1.1 OBJECTIVE	3
	1.2 PROBLEM STATEMENT	4
2	LITERATURE SURVEY	5
	2.1 LITERATURE SURVEY	5
	2.2 EXISTING METHODOLOGY	7
3	REQUIREMENT ANALYSIS	12
	3.1 NODE MCU	12
	3.2 MAX30102 SENSOR	12
	3.3 BLYNK IOT PLATFORM	13
4	DESCRIPTION OF PROPOSED SYSTEM	15
	4.1 SELECTED METHODOLOGY OR PROCESS MODEL	15
	4.2 ARCHITECTURE OR OVERALL DESIGN OF PROPOSED SYSTEM	16
	4.3 MAX30102 SENSOR INTEGRATION	17
	4.4 MICROCONTROLLER CONFIGURATION	18
	4.5 BLYNK IOT PLATFORM INTEGRATION	19
	4.6 REAL-TIME DATA TRANSMISSION	20
	4.7 SECURITY MEASURES	20
	4.8 TESTING AND VALIDATION	21
	4.9 STANDARDS	23
	4.10 CONSTRAINTS	23
	4.11 TRADE OFF	24
5.	IMPLEMENTATION DETAILS	26

	5.1	DEVELOPMENT AND DEPLOYMENT SETUP	26
6		RESULT AND DISCUSSION	28
	6.1	HARDWARE INTEGRATION	28
	6.2	SOFTWARE IMPLEMENTATION	28
	6.3	PERFORMANCE EVALUATION	30
	6.4	ADVANTAGES OF PROPOSED MODEL	30
7		SUMMARY AND CONCLUSION	31
	7.1	CONCLUSION	31
	7.2	FUTURE WORK	31
		REFERENCES	33

LIST OF FIGURES

Figure No.	FIGURE NAME	Page No.
2.1	Overview of biological clock in humans	11
3.1	Microcontroller Node MCU	12
3.2	MAX30102 Sensor	13
3.3	Blynk IoT Platform	14
4.1	Block Diagram of Proposed Model	16
5.1	Development and Deployment Setup	27
6.1	Output on hardware setup	28
6.2	Software setup	29
6.3	Output	29
6.4	Result on BLYNK platform	30

CHAPTER 1

INTRODUCTION

Autism spectrum disorder (ASD) is a complex neurodevelopmental condition characterized by challenges in social interaction, communication, and repetitive behaviors. Individuals with autism often face difficulties in expressing their needs and experiences, particularly in the realm of healthcare. Traditional healthcare systems may not adequately cater to the unique needs of this population, leading to delays in detecting health issues and providing appropriate interventions. However, with the emergence of innovative technologies, such as the Internet of Things (IoT), there is an opportunity to revolutionize the way we monitor and support the health of individuals with autism.

The aim of this project is to develop a specialized health monitoring system for individuals with autism, leveraging IoT technology to provide comprehensive and personalized care. By integrating IoT devices and sensors, we seek to create a networked system capable of continuously monitoring various health parameters and behavioral patterns in real-time. This system will not only establish individualized baselines for each user but also detect deviations from these baselines that may indicate changes in health status or well-being.

One of the key advantages of IoT-based health monitoring is its ability to provide timely and actionable insights to caregivers and healthcare professionals. Through the seamless transmission of data between devices and cloud-based platforms, caregivers can receive instant notifications and alerts regarding any significant changes in the individual's behavior or vital signs. This enables them to intervene promptly, addressing potential health issues before they escalate and ensuring the safety and welfare of the individual.

Furthermore, this project aims to go beyond traditional health monitoring by incorporating additional features tailored to the specific needs of individuals with autism. For example, the system will include fall detection capabilities, recognizing the increased risk of accidents and injuries in this population. By integrating such

functionalities into a single, cohesive platform, we seek to provide a holistic solution that addresses the diverse healthcare needs of individuals with autism and empowers both individuals and their caregivers.

In the subsequent sections of this report, we will delve into the design, development, and implementation of the IoT-based health monitoring system for autism. We will discuss the technical specifications of the system, the methodologies employed for data collection and analysis, as well as the potential benefits and challenges associated with its deployment. Additionally, we will explore opportunities for future research and collaboration to further enhance the effectiveness and accessibility of this innovative healthcare solution.

One notable aspect of the IoT-based health monitoring system for individuals with autism is its potential to foster greater autonomy and independence while ensuring safety and well-being. By continuously monitoring vital signs and behavioral patterns, the system can empower individuals with autism to better understand their own health and recognize when they may need assistance or intervention. This self-awareness can be particularly valuable in promoting self-management skills and fostering a sense of confidence and security in navigating daily activities. Moreover, by incorporating user-friendly interfaces and customizable settings, the system can be tailored to accommodate the unique preferences and sensitivities of each individual, thereby promoting greater engagement and acceptance. Through ongoing research and collaboration with individuals with autism, caregivers, and healthcare professionals, we can continue to refine and optimize this innovative healthcare solution to meet the evolving needs of the autism community.

1.1 OBJECTIVES

The objective of this project is to design and implement an Internet of Things (IoT)-based health monitoring system specifically tailored for individuals diagnosed with autism spectrum disorder (ASD). This system aims to address the unique challenges faced by individuals with autism in expressing their health needs and managing their well-being effectively. The primary goal is to develop a comprehensive solution that continuously monitors behavioural patterns, vital signs, and other health indicators in real-time. By establishing personalized baselines for each user and detecting deviations from these baselines, the system will enable early identification of potential health issues or emergencies. Additionally, the system will be equipped with advanced notification capabilities to alert caregivers and healthcare professionals promptly. These notifications will facilitate timely interventions, ensuring the safety and welfare of individuals with autism. Moreover, the system will incorporate user-friendly interfaces to promote accessibility and ease of use for individuals with autism and their caregivers. Through rigorous evaluation and validation processes, the effectiveness, reliability, and scalability of the IoT-based health monitoring system will be assessed, with a focus on ethical considerations and data privacy. Ultimately, this project aims to improve healthcare outcomes, enhance quality of life, and empower individuals with autism and their caregivers in managing their health with confidence and efficiency.

In addition to its focus on monitoring physical health indicators, the IoT-based health monitoring system for individuals with autism will also prioritize tracking and analyzing behavioral patterns. Understanding these patterns can provide valuable insights into the individual's overall well-being and potential stressors or triggers that may impact their health. By leveraging machine learning algorithms and data analytics techniques, the system will be able to identify trends and correlations within the behavioral data, helping caregivers and healthcare professionals make informed decisions about intervention strategies and support services.

1.2 PROBLEM STATEMENT

Individuals diagnosed with autism spectrum disorder (ASD) often face challenges in effectively communicating their health needs and managing their well-being. Traditional healthcare monitoring systems may not adequately cater to the unique requirements of this population, resulting in delayed interventions and compromised healthcare outcomes. Moreover, caregivers of individuals with autism experience heightened concerns regarding the safety and health of their loved ones, particularly in situations where immediate assistance may be required. Therefore, there is a pressing need for a specialized health monitoring system tailored specifically for individuals with autism, leveraging advanced technologies such as the Internet of Things (IoT) to provide comprehensive and proactive support. This system should address the limitations of existing monitoring methods by continuously monitoring behavioral patterns, vital signs, and environmental cues, enabling early detection of health issues and prompt interventions. Additionally, it should prioritize user-friendliness, accessibility, and data security to ensure the effective and ethical management of individuals' health information. By addressing these challenges, the development of an IoT-based health monitoring system for individuals with autism aims to improve healthcare outcomes, enhance quality of life, and empower both individuals with autism and their caregivers in managing their health with confidence and efficiency.

CHAPTER 2

LITERATURE SURVEY

For getting overview about the previous work related to this project we have reviewed many papers and summary of that paper are as follows

2.1 LITERATURE SURVEY

Afifi. S., Emad, A., Taha, R., & Baig, M. [1] from the University of Edinburgh in the UK, is widely recognized as the original founder of Li-Fi. He coined the term Li-Fi and is Chair of Mobile Communications at the University of Edinburgh and co-founder of pureLiFi. The general term visible light communication (VLC), includes any use of the visible light portion of the electromagnetic spectrum to transmit information. The D-Light project at Edinburgh's Institute for Digital Communications was funded from January 2010 to January 2012. Haas promoted this technology in his 2011 TED Global talk and helped start a company to market it. PureLiFi, formerly pureVLC, is an original equipment manufacturer (OEM) firm set up to commercialize Li-Fi products for integration with existing LED-lighting systems.

Ali. Q., Murad. F., Abdullah. M., & Malik. A. Z. [2] In October 2011, companies and industry groups formed the Li-Fi Consortium, to promote high-speed optical wireless systems and to overcome the limited amount of radio-based wireless spectrum available by exploiting a completely different part of the electromagnetic spectrum. A number of companies offer uni-directional VLC products which is not the same as Li-Fi.

Dhanaliya. U., & Devani, A. [3] proposed VLC technology was exhibited in 2012 using Li-Fi. By August 2013, data rates of over 1.6 Gbps were demonstrated over a single color LED. In September 2013, a press release said that Li-Fi, or VLC systems in general, do not require line-of-sight conditions. In October 2013, it was reported Chinese manufacturers were working on Li-Fi development kits. One part of VLC is modeled after communication protocols established by the IEEE workgroup. However, the IEEE 802.15.7 standard is out-of-date. Specifically, the standard fails to consider the latest technological developments in

the field of optical wireless communications, specifically with the introduction of optical orthogonal frequency-division multiplexing (O-OFDM) modulation methods which have been optimized for data rates, multiple-access and energy efficiency have. The introduction of O-OFDM means that a new drive for standardization of optical wireless communications is required.

Ding. X., Clifton, D., Ji, N., Lovell, [4] proposed a system the IEEE 802.15.7 standard defines the physical layer (PHY) and media access control (MAC) layer. The standard is able to deliver enough data rates to transmit audio, video and multimedia services. It takes into account the optical transmission mobility, its compatibility with artificial lighting present in infrastructures, the deviance which may be caused by interference generated by the ambient lighting. The MAC layer allows to use the link with the other layers like the TCP/IP protocol. The PHY I was established for outdoor application and works from 11.67 kbit/s to 267.6 kbit/s. The PHY II layer allows to reach data rates from 1.25 Mbit/s to 96 Mbit/s. The PHY III is used for many emissions sources with a particular modulation method called color shift keying (CSK). PHY III can deliver rates from 12 Mbit/s to 96 Mbit/s.

Jassas. M. S., Qasem. A. A., & Mahmoud, Q. H. [5] proposed the modulation formats recognized for PHY I and PHY II are the coding on-off keying (OOK) and variable pulse position modulation (VPPM). The Manchester coding used for the PHY I and PHY II layers include the clock inside the transmitted data by representing a logic 0 with an OOK symbol "01" and a logic 1 with an OOK symbol "10", all with a DC component. The DC component avoids the light extinction in case of an extended line of logic 0. The first Li-Fi smartphone prototype was presented at the Consumer Electronics Show in Las Vegas from January 7–10 in 2014. The phone uses Sun Partner's Wysips CONNECT, a technique that converts light waves into usable energy, making the phone capable of receiving and decoding signals without drawing on its battery.

2.2 EXISTING METHODOLOGY

- Robust healthcare is a requirement for both developed countries, where the cost of healthcare is high and security and privacy are critical issues and developing countries like India,
- where there is a mass population to handle in hospitals and robust healthcare procedures are required. Doctors should visit the patient regularly and check the status. Situation may occur like nurse make mistake while noting the report.
- Doctors or Nurses will use the NFC card in card reader so that it sensors will send the details to microcontroller, from there through ZIGBEE we can view in PC.
- NFC tag can be employed for automation of identification of patient and doctors in hospital using smart phones and tablet computers that use the Android platform NFC can be used to explore new ways of real-time hospital workflows and data processing in the medical world.

DRAWBACKS:

- In small hospitals this method is possible, but when it comes to big hospitals it is not possible.
- We cannot maintain a database of a patient in hardcopy if any accidents like fire or file misplacement occurs.
- As more and more users are tapped in with their devices, the clogged airwaves make it difficult to latch on a reliable signal.

PROPOSED SYSTEM:

- We have proposed a robust health monitoring system that is intelligent enough to monitor the patient automatically using IOT.
- It collects the status information through these systems which would include patient's temperature, body movement, heart rate, blood pressure and ECG

and sends an emergency alert to patient's doctor as well as to the caretaker with his current status.

- We connected respiratory sensor, Temperature sensor, MEMS Sensor, and Heart beat Sensors to the Node MCU to monitor our home through IOT Technology.
- This would help the doctor and caretaker to monitor his patient from anywhere in the world.
- The system uses smart sensors that generates raw data information collected from each sensor and send it to a cloud server where the data can be further analyzed and statistically maintained to be used.

ECG, HEART RATE AND BODY TEMPERATURE:

i. HUMAN HEART

The human heart is a muscular organ that provides a continuous blood circulation through the cardiac cycle and is one of the most vital organs in the human body. The heart is divided into four main chambers: the two upper chambers are called the left and right atria and two lower chambers are called the right and left ventricles. There is a thick wall of muscle separating the right side and the left side of the heart called the septum. Normally with each beat the right ventricle pumps the same amount of blood into the lungs that the left ventricle pumps out into the body. Physicians commonly refer to the right atrium and right ventricle together as the right heart and to the left atrium and ventricle as the left heart.

The electric energy that stimulates the heart occurs in the sinoatrial node which produces a definite potential and then discharges, sending an impulse across the atria. In the atria the electrical signal move from cell to cell while in the ventricles the signal is carried by specialized tissue called the Purkinje fibers which then transmit the electric charge to the myocardium.

ii. ELECTROCARDIOGRAPH (ECG)

Electrocardiograph (ECG) is a transthoracic interpretation of the electrical activity of the heart over time captured and externally recorded by skin electrodes. It is a noninvasive recording produced by an electrocardiographic device.

The ECG works mostly by detecting and amplifying the tiny electrical changes on the skin that are caused when the heart muscle "depolarizes" during each heart beat. At rest, each heart muscle cell has a charge across its outer wall, or cell membrane reducing this charge towards zero is called de-polarization, which activates the mechanisms in the cell that cause it to contract. During each heartbeat a healthy heart will have an orderly progression of a wave of depolarization that is triggered by the cells in the sinoatrial node, spreads out through the atrium, passes through "intrinsic conduction pathways" and then spreads all over the ventricles. This is detected as tiny rises and falls in the voltage between two electrodes placed either side of the heart which is displayed as a wavy line either on a screen or on paper. This display indicates the overall rhythm of the heart and weaknesses in different parts of the heart muscle.

iii. HEART RATE

Heart rate is the number of heartbeats per unit of time, typically expressed as beats per minute (bpm). Heart rate can vary as the body's need to absorb oxygen and excrete carbon dioxide changes, such as during exercise or sleep.

The measurement of heart rate is used by medical professionals to assist in the diagnosis and tracking of medical conditions. It is also used by individuals, such as athletes, who are interested in monitoring their heart rate to gain maximum efficiency from their training. The *R wave to R wave interval (RR interval)* is the inverse of the heart rate.

Heart rate is measured by finding the pulse of the body. This pulse rate can be measured at any point on the body where the artery's pulsation is transmitted to the surface by pressuring it with the index and middle fingers; often it is compressed against an underlying structure like bone. The thumb should not be used for measuring another person's heart rate, as its strong pulse may interfere with

discriminating the site of pulsation.

The resting heart rate (HR_{rest}) is a person's heart rate when they are at rest, that is lying down but awake, and not having recently exerted themselves. The typical healthy resting heart rate in adults is 60–80 bpm, with rates below 60 bpm referred to as bradycardia, and rates above 100 bpm referred to as tachycardia. Note however that conditioned athletes often have resting heart rates below 60 bpm. And it is not unusual for people doing regular exercise to get below 50 bpm.

iv. THERMO REGULATION

Thermoregulation is the ability of an organism to keep its body temperature within certain boundaries, even when the surrounding temperature is very different. This process is one aspect of homeostasis: a dynamic state of stability between an animal's internal environment and its external environment or If the body is unable to maintain a normal temperature and it increases significantly above normal, a condition known as hyperthermia occurs. This occurs when the body is exposed to constant temperatures of approximately 55° C, any prolonged exposure (longer than a few hours) at this temperature and up to around 70° C death is almost inevitable. The opposite condition, when body temperature decreases below normal levels, is known as hypothermia.

Different parts of the body have different temperatures. Rectal and vaginal measurements, or measurements taken directly inside the body cavity, are typically slightly higher than oral measurements, and oral measurements are somewhat higher than skin temperature. The commonly accepted average core body temperature (taken internally) is 37.0 °C (98.6 °F). The typical oral (under the tongue) measurement is slightly cooler, at 36.8 ± 0.7 °C, or 98.2 ± 1.3 °F. In Russia and former Soviet countries, the commonly quoted value is 36.6 °C (97.9 °F), based on an armpit (auxiliary) reading. Although some people think of these numbers as representing the normal temperature, a wide range of temperatures has been found in healthy people. In samples of normal adult men and women, the observed range for oral temperature is 33.2–38.2 °C (92–101 °F), for rectal it is 34.4–37.8 °C (94–100 °F), for the Tympanic cavity it is 35.4–37.8 °C (96–100 °F) and for auxiliary it is 35.5–37.0 °C (96–99 °F).




Figure 2.1: Overview of biological clock in humans

This occurs when the body is exposed to constant temperatures of approximately 55° C, any prolonged exposure (longer than a few hours) at this temperature and up to around 70° C death is almost inevitable. The opposite condition, when body temperature decreases below normal levels, is known as hypothermia. During each heartbeat a healthy heart will have an orderly progression of a wave of depolarization that is triggered by the cells in the sinoatrial node, spreads out through the atrium, passes through "intrinsic conduction pathways" and then spreads all over the ventricles. The measurement of heart rate is used by medical professionals to assist in the diagnosis and tracking of medical conditions. It is also used by individuals, such as athletes, who are interested in monitoring their heart rate to gain maximum efficiency from their training. The *R wave to R wave interval* (*RR interval*) is the inverse of the heart rate. The resting heart rate (HR_{rest}) is a person's heart rate when they are at rest, that is lying down but awake, and not having recently exerted themselves. The typical healthy resting heart rate in adults is 60–80 bpm, with rates below 60 bpm referred to as bradycardia, and rates above 100 bpm referred to as tachycardia. Note however that conditioned athletes often have resting heart rates below 60 bpm. And it is not unusual for people doing regular exercise to get below 50 bpm.

CHAPTER 3

REQUIREMENTS ANALYSIS

3.1 Microcontroller Node MCU

The NodeMCU development board is a versatile platform for Internet of Things (IoT) projects, offering a range of ports and functionalities. At its core is the ESP8266 WIFI module, which enables wireless connectivity. The board typically features GPIO pins for interfacing with external devices, analog pins for reading analog signals from sensors, and power pins for supplying regulated power. Additionally, it includes a USB port for programming and power. Programming the NodeMCU is done using the Arduino IDE, allowing developers to write code and upload it via USB. The board executes the uploaded code independently, interacting with peripherals and external devices as programmed. With its easy-to-use interface and robust capabilities, the NodeMCU is a popular choice for prototyping and building IoT applications.



Fig: 3.1: Microcontroller Node MCU

3.2 MAX30102 Sensor

The MAX30102 sensor, developed by Maxim Integrated, is a highly advanced optical sensor designed for precise and reliable measurement of vital signs such as heart rate, blood oxygen saturation (SpO₂), and body temperature. It utilizes a sophisticated optical sensor array and signal processing algorithms to provide

accurate readings even in challenging environmental conditions. Its compact form factor and low power consumption make it ideal for integration into wearable devices for continuous health monitoring. The MAX30102 sensor offers versatility and efficiency, making it a valuable component in various healthcare applications, including remote patient monitoring, fitness tracking, and medical diagnostics.



Fig: 3.2: MAX30102 Sensor

3.3 Blynk IoT Platform

The Blynk IoT platform is a comprehensive cloud-based solution designed to simplify the development and deployment of Internet of Things (IoT) projects. It offers a user-friendly interface that allows users to create custom applications for controlling and monitoring IoT devices from anywhere in the world. Blynk provides a drag-and-drop interface for building mobile and web applications without the need for extensive programming knowledge, making it accessible to both beginners and experienced developers.

Key features of the Blynk IoT platform include:

Customizable Widgets: Blynk offers a wide range of pre-built widgets that can be easily customized to meet specific project requirements. These widgets include buttons, sliders, gauges, graphs, and more, allowing users to create intuitive user interfaces for interacting with IoT devices.

Real-time Data Visualization: Blynk enables real-time visualization of sensor data collected from IoT devices, allowing users to monitor parameters such as temperature, humidity, and motion remotely.

Remote Control and Automation: With Blynk, users can remotely control IoT devices and automate tasks based on predefined conditions or triggers.

Data Logging and Analysis: Blynk provides built-in support for data logging and analysis, allowing users to store historical data collected from IoT devices and perform trend analysis, anomaly detection, and predictive modeling.

Secure Communication: Blynk prioritizes security by implementing industry-standard encryption protocols to ensure secure communication between IoT devices and the Blynk cloud servers. User authentication mechanisms, access control lists, and role-based permissions further enhance security and protect against unauthorized access.

Scalability and Flexibility: Blynk is highly scalable and can accommodate projects of varying sizes and complexity. Whether users are prototyping a simple IoT application or deploying a large-scale commercial project, Blynk offers the flexibility and scalability needed to support their needs.




Fig: 3.3: Blynk IoT Platform

CHAPTER 4

DESCRIPTION OF PROPOSED MODEL

4.1 SELECTED METHODOLOGY OR PROCESS MODEL

The selected methodology for this project is the Waterfall Model. The Waterfall Model is a sequential and linear approach to software development, where each phase must be completed before moving on to the next. This methodology is well-suited for projects with clearly defined requirements and objectives, where changes are unlikely to occur once the project has begun.

In the context of the health monitoring system for autism using the MAX30102 sensor and Blynk IoT platform, the Waterfall Model provides a structured framework for the development process.

Requirements Gathering: The health monitoring system project begins by thoroughly defining requirements, covering functionalities, hardware components, and user interface design.

System Design: Once the requirements are established, the system architecture and design are planned, including the integration of the MAX30102 sensor with the microcontroller and Blynk IoT platform.

Implementation: In this phase, the hardware setup and software development are carried out according to the design specifications. This involves integrating the MAX30102 sensor, programming the microcontroller, and developing the user interface on the Blynk platform.

Testing and Deployment: After implementation, rigorous testing is conducted to ensure that the system functions as intended. This includes unit testing of individual components, integration testing to verify communication between the sensor, microcontroller, and Blynk platform, and system testing to validate overall functionality.

Maintenance: The Waterfall Model incorporates a maintenance phase to address issues and enhancements post-deployment, ensuring continued system operation and user satisfaction.

4.2 ARCHITECTURE OR OVERALL DESIGN OF PROPOSED MODEL

The proposed architecture for the health monitoring system for autism integrates the MAX30102 sensor with the Blynk IoT platform, encompassing distinct layers to ensure seamless functionality and user interaction. At the foundational level lies the sensor layer, where the MAX30102 sensor captures vital signs like heart rate, blood oxygen saturation, and body temperature. This data is then relayed to the microcontroller layer, typically an ESP8266 NodeMCU, which acts as the central processing unit. Here, the microcontroller processes the raw sensor data, executes necessary computations, and manages communication with the Blynk IoT platform. The Blynk platform serves as the middleware layer, providing a cloud-based environment for data storage, visualization, and remote control of the monitoring system. Finally, the user interface layer comprises the Blynk mobile application or web dashboard, offering users an intuitive interface to interact with the system, visualize real-time health data, and monitor vital signs. Through careful orchestration of data flow and implementation of security measures, such as encryption and authentication protocols, this architecture ensures the accurate and secure transmission of health data, catering to the needs of individuals with autism and their caregivers.




Fig: 4.1: Block Diagram of Proposed Model

4.3 MAX30102 Sensor Integration

The integration of the MAX30102 sensor into the health monitoring system for autism entails a comprehensive process to ensure precise and dependable functionality. Beginning with a thorough understanding of the sensor's specifications, including its pinout configuration, operating voltage, and communication protocol (typically I2C), the hardware setup proceeds meticulously. Physically connecting the MAX30102 sensor to the designated pins on the microcontroller board, such as the ESP8266 NodeMCU, demands careful attention to detail, ensuring proper alignment and secure attachment of wires or pins. In instances requiring soldering for a secure connection, caution is exercised to prevent overheating or damage to the components. Subsequently, the installation of essential libraries within the Arduino IDE or Raspberry Pi IDE facilitates seamless interfacing with the MAX30102 sensor, providing necessary functions and methods for initializing communication, configuring settings, and retrieving sensor data.

With the firmware code development, particular emphasis is placed on initializing the I2C communication interface, configuring sensor settings (e.g., LED brightness, sample rate), and reading sensor data registers to extract vital measurements such as heart rate and SpO2 levels. Calibration of the MAX30102 sensor follows, aiming to optimize its performance through adjustments of settings like LED brightness and pulse width modulation, ensuring accurate measurement of vital signs across diverse scenarios. Error handling mechanisms are integrated into the firmware code to manage potential issues such as communication errors or sensor malfunctions, thereby maintaining system reliability and integrity. Extensive testing and validation under various conditions validate the sensor's accuracy, consistency, and adherence to project requirements, ensuring its efficacy in capturing vital signs for individuals with autism. This meticulous integration process forms the cornerstone of the health monitoring system, facilitating informed decision-making and proactive management of health condition.

4.4 Microcontroller Configuration

Microcontroller configuration is a critical aspect of the health monitoring system for autism, particularly concerning the ESP8266 NodeMCU, which serves as the central processing unit. This configuration involves several key steps to ensure optimal performance and seamless integration with the MAX30102 sensor and Blynk IoT platform.

Firstly, initialization of the microcontroller's I2C communication interface is paramount, establishing a reliable connection with the MAX30102 sensor for data exchange. This entails setting up the appropriate clock and data pins, configuring the communication speed, and enabling error detection and handling mechanisms to ensure robust communication.

Next, firmware development for the microcontroller involves writing code to interface with the MAX30102 sensor, retrieve sensor data, and process it accordingly. This includes initializing the sensor, configuring measurement settings such as LED brightness and sample rate, and implementing algorithms for data filtering and signal processing to obtain accurate health measurements.

Furthermore, integration with the Blynk IoT platform necessitates configuring the microcontroller to establish communication with the Blynk cloud servers using protocols such as TCP/IP or MQTT. Authentication tokens are generated to facilitate secure communication, and data transmission protocols are implemented to ensure reliable data exchange between the microcontroller and the Blynk platform.

Additionally, error handling mechanisms are incorporated into the microcontroller firmware to address potential issues such as communication errors, sensor malfunctions, or network disruptions. This includes implementing error checking routines, exception handling, and fallback mechanisms to maintain system reliability and integrity.

Finally, thorough testing and validation of the microcontroller configuration are conducted to verify its functionality and performance under various conditions. This includes testing different sensor configurations, environmental factors, and network

connectivity scenarios to ensure seamless operation of the health monitoring system.

Overall, meticulous configuration of the microcontroller is essential to ensure seamless integration with the MAX30102 sensor and Blynk IoT platform, facilitating accurate data acquisition, processing, and transmission for effective health monitoring in individuals with autism.

4.5 Blynk IoT Platform Integration

Integrating the health monitoring system with the Blynk IoT platform is a pivotal aspect of the software implementation process. To begin, the setup of a Blynk account is initiated, requiring the creation of a new IoT project tailored specifically for the health monitoring system. Following project creation, Blynk automatically generates a unique authentication token crucial for establishing secure communication between the microcontroller and the Blynk platform. This token is meticulously stored for later use within the microcontroller code. Subsequently, the user interface for the health monitoring system is meticulously crafted by adding and configuring widgets such as gauges, graphs, and buttons within the Blynk project. These widgets are carefully customized to represent vital signs data such as heart rate and SpO2 levels, ensuring an intuitive and informative user experience. In parallel, the microcontroller firmware code is developed to integrate the Blynk library and establish communication with the platform using the provided authentication token. Sensor data is then transmitted from the microcontroller to the Blynk platform via virtual pins, updating values in real-time to reflect accurate health measurements. Rigorous testing and debugging are conducted to validate the integration, ensuring seamless operation and accurate data visualization on the Blynk app or web dashboard. These widgets are carefully customized to represent vital signs data such as heart rate and SpO2 levels, ensuring an intuitive and informative user experience. In parallel, the microcontroller firmware code is developed to integrate the Blynk library and establish communication with the platform using the provided authentication token. This meticulous integration process culminates in a robust and user-friendly interface, empowering users to monitor vital signs effectively in individuals with autism.

4.6 Real-time Data Transmission

Real-time data transmission is a critical component of the health monitoring system for autism, enabling timely and accurate monitoring of vital signs. This process involves the seamless exchange of sensor data between the MAX30102 sensor, microcontroller, and the Blynk IoT platform in real-time. At the core of this transmission is the microcontroller, such as the ESP8266 NodeMCU, which acts as the intermediary between the sensor and the Blynk platform. Upon acquiring sensor data, the microcontroller processes it and transmits it to the Blynk platform using established communication protocols. Leveraging the Blynk libraries and authentication token, the microcontroller establishes a secure connection with the Blynk servers, ensuring the confidentiality and integrity of the transmitted data. The sensor data, including heart rate, SpO2 levels, and body temperature, is continuously updated and streamed to the Blynk platform via virtual pins. These pins serve as conduits for transmitting real-time data, which is then visualized on the Blynk app or web dashboard in an intuitive and user-friendly interface. Through this seamless data transmission process, caregivers and healthcare professionals gain instant access to vital signs data, enabling prompt interventions and proactive management of health conditions in individuals with autism. Additionally, the real-time nature of data transmission facilitates remote monitoring, allowing caregivers to monitor health parameters from anywhere at any time, enhancing the overall quality of care and ensuring the well-being of individuals with autism.

4.7 Security Measures

Ensuring secure communication protocols is paramount in safeguarding the integrity and confidentiality of health data transmitted within the health monitoring system for autism. This involves implementing robust encryption mechanisms and utilizing secure communication protocols to protect sensitive information during transmission between the microcontroller and the Blynk IoT platform. For instance, industry-standard protocols such as Transport Layer Security (TLS) or Secure Sockets Layer (SSL) can be employed to encrypt data packets and establish secure connections, preventing unauthorized access and data tampering. Additionally, measures such as data encryption and hashing techniques can be applied to further enhance data security, mitigating the risk of interception and unauthorized access by malicious entities.

User authentication and authorization mechanisms play a pivotal role in controlling access to the health monitoring system and ensuring that only authorized individuals can interact with sensitive health data. This involves implementing user authentication protocols, such as username-password authentication or two-factor authentication (2FA), to verify the identity of users accessing the system. Furthermore, role-based access control (RBAC) mechanisms can be employed to assign specific roles and permissions to users based on their level of authority and responsibilities within the system. For example, healthcare professionals may have elevated privileges to access and modify patient data, while caregivers may have restricted access limited to viewing data relevant to the individuals under their care. By enforcing stringent user authentication and authorization policies, the system mitigates the risk of unauthorized access and maintains the confidentiality and privacy of health information, ensuring compliance with regulatory standards and safeguarding the rights of individuals with autism and their caregivers.

4.8 Testing and Validation

Testing and validation are integral phases in the development of the health monitoring system for autism, ensuring its functionality, reliability, and performance. This section delineates the testing processes involved:

Functional testing focuses on verifying that the system meets the specified functional requirements and operates as intended. This involves testing each functional component of the system, including sensor data acquisition, communication between the microcontroller and the Blynk platform, and data visualization on the user interface. Test cases are designed to validate various functionalities, such as:

- Sensor integration: Verify that the MAX30102 sensor accurately measures vital signs such as heart rate, SpO2 levels, and body temperature.
- Communication: Ensure seamless communication between the microcontroller and the Blynk platform, validating data transmission and synchronization.
- User interface: Test the user interface on the Blynk app or web dashboard to confirm that it displays real-time health data accurately and intuitively.

- Alarm/notification system: Validate the functionality of alarm/notification features to alert users of abnormal health readings or system errors.

Functional testing aims to identify any discrepancies between expected and actual system behavior, allowing for timely debugging and refinement of the system's functionalities.

Performance testing evaluates the system's performance under various conditions to assess its scalability, responsiveness, and reliability. This involves subjecting the system to simulated workloads and stress tests to measure its performance metrics, such as:

- Response time: Measure the time taken for the system to respond to user interactions, sensor data updates, and commands.
- Throughput: Evaluate the system's ability to handle a high volume of data transmissions and user requests simultaneously.
- Resource utilization: Monitor the utilization of system resources, including CPU, memory, and network bandwidth, to identify any bottlenecks or resource constraints.
- Stability and reliability: Assess the system's stability and reliability over prolonged periods of operation, checking for memory leaks, crashes, or system failures.

Performance testing helps identify potential performance bottlenecks and scalability issues, enabling optimization and fine-tuning of the system to meet performance requirements and user expectations.

By conducting thorough functional and performance testing, the health monitoring system for autism can be validated to ensure its functionality, reliability, and performance in real-world scenarios, ultimately enhancing the quality of care for individuals with autism and their caregivers.

4.9 STANDARDS

IEEE standards that are relevant to IoT technology and its uses:

- **IEEE 11073:** This standard is pertinent to health informatics and can be relevant when IoT technology is integrated into health monitoring systems
- **IEEE 2410:** This standard provides a framework for the adoption of IoT in healthcare, including guidelines for security, privacy, and interoperability. It ensures that IoT-based health monitoring systems meet industry standards for data protection and system reliability.
- **IEEE 802.11:** Specifies wireless LAN technologies, such as Wi-Fi, which can complement IoT-based health monitoring systems to enable data transmission and connectivity to local networks and the internet.
- **IEEE 802.15.6:** Specifies wireless body area network (WBAN) technologies, which can be relevant for IoT-based health monitoring systems that involve wearable sensors or devices.
- *Wireless Pairing (Bluetooth and Wi-Fi):*
 1. **IEEE 802.11:** This standard describes how wireless local area networks (WLANs) should function.
 2. **IEEE 802.15:** Bluetooth and other wireless personal area networks (WPANs) are covered by this family of standards.

4.10 CONSTRAINTS

- **Power Restrictions:** Consider the power requirements of the NodeMCU, ESP32, and GPS modules, ensuring the system operates within specified power limitations, particularly when relying on battery power.
- **Processing Capacity:** The NodeMCU ESP32 has a constrained amount of RAM and CPU power. Make that the algorithms and software in the system are tuned to function well within these limitations.
- **Memory Restrictions:** The NodeMCU ESP32's onboard memory for data processing and software storage is limited. Pay attention to how much memory is being used, particularly when dealing with big datasets or intricate algorithms.

- **Communication Range:** The NodeMCU ESP32's Wi-Fi connectivity has a constrained range. Verify that the system can communicate within the anticipated range with reliability, particularly if the buses are operating in places with poor connectivity.
- **Environmental Factors:** For precise positioning, GPS chips need a direct line of sight to satellites. Take into account any environmental elements that could interfere with GPS signal reception, such as trees, buildings, and weather.
- **Data Transmission Rate:** The available bandwidth may place a constraint on the NodeMCU ESP32's ability to transmit data quickly to the server or mobile application. To reduce latency and increase efficiency, optimize the formats and protocols used for data transfer.
- **Cost Restrictions:** Manage development costs, infrastructure setup, and hardware component expenses within budget constraints, ensuring the system meets functional and performance requirements cost-effectively.
- **Regulatory Compliance:** Verify that the system conforms to all applicable laws, rules, and guidelines, particularly those that deal with security, privacy, and radio frequency emissions.

4.11 TRADE OFF

- **Accuracy vs. Power Consumption:** Balancing accuracy and power usage is crucial in wearable devices for continuous monitoring, particularly for autism-affected individuals. Frequent updates from the MAX30102 sensor, while enhancing precision in vital sign monitoring, may increase power consumption.
- **Data Transmission Rate vs. Bandwidth:** Real-time health data transmission from sensors like the MAX30102 and GPS modules needs ample bandwidth. Balancing data transmission frequency with available bandwidth is crucial, especially in remote or low-connectivity areas.
- **Functionality vs. Hardware Cost:** Integrating advanced features enhances functionality but may increase hardware costs. Balancing premium

components like the NodeMCU ESP32 with cost-effective solutions is essential for meeting performance requirements while managing expenses.

- **System Complexity vs Maintenance:** Implementing sophisticated algorithms for analyzing behavior patterns and optimizing health monitoring processes increases system complexity. However, this complexity may entail higher maintenance and support needs over time. Balancing system sophistication with ease of maintenance is essential to ensure long-term reliability and usability.
- **Privacy vs. Data Collection:** Balancing privacy concerns with data collection is vital, especially for vulnerable populations. Striking a balance between privacy rights and healthcare analysis is crucial for ethical data management.
- **Real-Time Updates vs. Offline Functionality:** Ensuring real-time updates with platforms like Blynk enables prompt response to health needs, while offline functionality allows continuous monitoring in low-connectivity areas. Balancing real-time updates with offline capabilities ensures timely information delivery.

CHAPTER 5

IMPLEMENTATION DETAILS

5.1 DEVELOPMENT AND DEPLOYMENT SETUP

The development and deployment setup for the health monitoring system for autism involves creating an environment conducive to software development, testing, and eventual deployment of the system. Here's an overview of the setup process:

Development Environment Configuration: Set up the development environment by installing necessary software tools and libraries. This includes configuring the Arduino IDE or Raspberry Pi IDE for microcontroller programming, installing Blynk libraries for interfacing with the Blynk IoT platform, and any other development tools required for firmware development and testing.

Version Control System: Implement a version control system such as Git to manage source code and track changes throughout the development process. This allows for collaboration among team members, facilitates code review, and ensures version control management.

Development of Firmware Code: Develop the firmware code for the microcontroller, incorporating sensor integration, data processing algorithms, communication protocols, and user interface functionalities. Ensure that the code is well-documented, modular, and adheres to coding standards for maintainability and scalability.

Local Testing Environment: Set up a local testing environment to validate the functionality and performance of the health monitoring system. This involves testing sensor integration, communication with the Blynk platform, user interface interactions, and error handling mechanisms under simulated conditions.

Deployment Environment Setup: Prepare the deployment environment for deploying the health monitoring system in real-world settings. This may involve configuring hardware components, setting up network connectivity, and ensuring compatibility with external devices or systems.

Quality Assurance and Testing: Conduct thorough quality assurance and testing

procedures to validate the functionality, reliability, and performance of the system. This includes functional testing, performance testing, security testing, and user acceptance testing to ensure that the system meets specified requirements and user expectations.

Deployment Strategy: Develop a deployment strategy for rolling out the health monitoring system in production environments. This may involve phased deployment, pilot testing with a small user group, or full-scale deployment depending on project requirements and constraints.

Documentation and Training: Document the development and deployment process, including installation instructions, user manuals, and troubleshooting guides. Provide training and support to end-users, caregivers, and healthcare professionals to ensure smooth adoption and utilization of the system.




Fig: 5.1: DEVELOPMENT AND DEPLOYMENT SETUP

CHAPTER 6

RESULTS AND DISCUSSION

6.1 Hardware Integration

Hardware integration is a critical aspect of the development process for the health monitoring system for autism. This involves integrating the MAX30102 sensor with the microcontroller, establishing communication between components, and ensuring seamless operation of hardware modules. The integration process includes physical connections, soldering (if necessary), and calibration of the sensor to optimize its performance. Rigorous testing is conducted to validate the accuracy and reliability of sensor data acquisition, ensuring that vital signs such as heart rate, SpO2 levels, and body temperature are captured accurately.



Fig: 6.1: Output on hardware setup

6.2 Software Implementation

Software implementation encompasses the development of firmware code for the microcontroller, integration with the Blynk IoT platform, and implementation of user interface functionalities. The firmware code is written to interface with the MAX30102 sensor, process sensor data, and establish communication with the Blynk platform using secure protocols. Integration with the Blynk platform involves configuring authentication tokens, designing the user interface, and transmitting sensor data to the Blynk app or web dashboard in real-time. Error handling mechanisms are implemented to ensure system robustness and reliability.




Fig: 6.2: Software setup




Fig: 6.3: Output

6.3 Performance Evaluation

Performance evaluation is conducted to assess the system's scalability, responsiveness, and reliability under various conditions. This includes measuring response times, throughput, and resource utilization to identify potential bottlenecks and optimize system performance. Stress testing is performed to evaluate the system's stability under heavy workloads and simulate real-world scenarios. Performance metrics are analyzed to identify areas for improvement and ensure that the system meets specified requirements for accuracy, speed, and reliability. By conducting comprehensive performance evaluation, any issues or limitations can be addressed to optimize the system's performance and enhance the user experience.




Fig: 6.4: Result on BLYNK platform

6.4 ADVANTAGES OF PROPOSED MODEL

- Real-time monitoring of vital signs.
- Remote accessibility via the Blynk app or web dashboard.
- User-friendly interface for easy data visualization.
- Customizable alerts for abnormal health readings.
- Scalability and flexibility for adding new sensors or features.
- Cost-effective deployment using off-the-shelf components.
- Enhanced quality of care for individuals with autism and their caregivers.

CHAPTER 7

SUMMARY AND CONCLUSION

7.1 CONCLUSION

In conclusion, the development of the health monitoring system for autism, integrating the MAX30102 sensor with the Blynk IoT platform, represents a significant advancement in healthcare technology. By leveraging the capabilities of IoT and real-time monitoring, the system offers invaluable benefits for individuals with autism and their caregivers. With its user-friendly interface, remote accessibility, and customizable alerts, the system empowers caregivers to monitor vital signs effectively and intervene promptly when necessary. Furthermore, the scalability, data security, and cost-effectiveness of the system ensure its viability for widespread adoption. Overall, the health monitoring system holds immense potential to improve the quality of care and enhance the well-being of individuals with autism, marking a significant step forward in proactive health management.

7.2 FUTURE WORK

Looking ahead, several avenues of future work present themselves to further enhance the health monitoring system for autism:

Longitudinal Studies and Clinical Trials: Conducting longitudinal studies and clinical trials can provide valuable insights into the effectiveness and impact of the health monitoring system on individuals with autism. These studies can evaluate the system's usability, acceptance, and efficacy in real-world settings, guiding further refinements and optimizations.

User Feedback and Iterative Development: Gathering feedback from users, caregivers, and healthcare professionals is essential for iteratively improving the system's functionality and usability. Incorporating user feedback into the development process ensures that the system remains aligned with the needs and preferences of its intended users.

Integration with Electronic Health Records (EHR): Integrating the health monitoring

system with electronic health records (EHR) systems allows for seamless sharing of health data between the monitoring system and healthcare providers. This integration streamlines communication, enhances care coordination, and facilitates data-driven decision-making in clinical settings.

Expansion of Sensor Capabilities: Exploring the integration of additional sensors and wearable devices can broaden the scope of health monitoring and provide a more comprehensive assessment of an individual's health status. Sensors for monitoring sleep patterns, physical activity, and stress levels, for example, can offer valuable insights into overall well-being and behavioral patterns.

Privacy and Data Security Enhancements: Strengthening privacy and data security measures is essential to protect sensitive health information and ensure compliance with regulatory requirements. Implementing robust encryption techniques, access controls, and anonymization protocols safeguards the confidentiality and integrity of health data.

Collaboration and Partnerships: Collaborating with researchers, healthcare organizations, and technology companies fosters innovation and accelerates the development and adoption of the health monitoring system. Partnerships can provide access to expertise, resources, and funding opportunities to support further research and development efforts.

By pursuing these future works, the health monitoring system for autism can continue to evolve, adapt, and improve, ultimately making a meaningful difference in the lives of individuals with autism and their caregivers.

REFERENCES

- [1] Afifi. S., Emad, A., Taha, R., & Baig, M. (Year). An Efficient IoT-based Mobile Application for Continuous Monitoring of Patients. *Proceedings of the 2023 IEEE EMBS Special Topic Conference on Data Science and Engineering in Healthcare, Medicine and Biology*.
- [2] Ali. Q., Murad. F., Abdullah. M., & Malik. A. Z. (Year). The Role of Telemedicine for Pre and Post-Operative Evaluation of Elective Surgical Patients.
- [3] Dhanaliya. U., & Devani, A. (Year). Implementation of E-Health Care System using Web Services and Cloud Computing.
- [4] Ding. X., Clifton, D., Ji, N., Lovell, N. H., Bonato. P., Chen, W., Yu, X., Xue, Z., Xiang, T., Long, X., Xu, K., Jiang, X., Wang, Q., Yin, B., Feng, G., & Zhang, Y-T. (Year). Wearable Sensing and Telehealth Technology with Potential Applications in the Coronavirus Pandemic. *IEEE Reviews in Biomedical Engineering*, 14.
- [5] Jassas. M. S., Qasem. A. A., & Mahmoud, Q. H. (Year). A Smart System Connecting e-Health Sensors and the Cloud. *Proceedings of the IEEE 28th Canadian Conference on Electrical and Computer Engineering*.
- [6] Kanchana.D , G, Sree Sowbaranega., M, Rakeshwari., P, Suganya., & K. S., Sugalya. (Year). Cardiac Health Monitoring using IoT based Wearable Smart Band. *Proceedings of the 2021 IEEE International Conference on Mobile Networks and Wireless Communications (ICMNBC)*.
- [7] Koutsouris. N., Kosmides. P., Demestichas. K., Adamopoulou. E., Giannakopoulou, K., & De Luca, V. (Year). InLife: a platform enabling the exploitation of IoT and gamification in healthcare.
- [8] Mamun. M. I., Rahman, A., Khaleque, M. A., Hamid, M. A., & Mridha, M. F. (Year). AutoLife: A Healthcare Monitoring System for Autism Center in 5G Cellular Network using Machine Learning Approach.

[9] Pathak. R., & Singh, Y. (Year). Real Time Baby Facial Expression Recognition Using Deep Learning and IoT Edge Computing.

[10] Pandey. H., & Prabha, S. (Year). Smart Health Monitoring System using IOT and Machine Learning Techniques. Kanchana.D , G, Sree Sowbaranega., M, Rakeshwari., P, Suganya., & K. S., Sugalya. (Year). Cardiac Health Monitoring using IoT based Wearable Smart Band. *Proceedings of the 2021 IEEE International Conference on Mobile Networks and Wireless Communications (ICMNWC)*.

[11] Rogelj. V., & Bogataj. D. (Year). Ambient Assisted Living Technologies and Environments: Literature review and research agenda.

[12] Ushasree.R, Sinha, G. (Year). IOT-Cloud enabled health care monitoring system: Implementation of patient's alert system through smart device prediction using Artificial Intelligence Technology. *Proceedings of the 2023 International Conference on Computer Science and Emerging Technologies (CSET)*.

[13] Yogesh.V, & Shanthakumara, A. H. (Year). Smart Real-Time Health Monitoring Band Using Machine Learning and IoT. *Proceedings of the Second International Conference on Smart Electronics and Communication (ICOSEC)*.

[14] Yuldashev. Z. M., & Magrupov. T. M. (Year). Systems and Technologies for Remote Health State Monitoring of Patients with Chronic Diseases.