

LOW COST NON INVASIVE ANALYSIS FOR MEASURING FAT AND WATER CONTENT OF THE BODY

Submitted in partial fulfillment of the requirements for the award of Bachelor of Technology degree in Biomedical Engineering

By

MOHAMED AZARUDEEN.S (40240019)

JEY SURYA.S (40240013)



**DEPARTMENT OF BIOMEDICAL ENGINEERING
SCHOOL OF BIO AND CHEMICAL ENGINEERING**

SATHYABAMA

**INSTITUTE OF SCIENCE AND TECHNOLOGY
(DEEMED TO BE UNIVERSITY)**

Accredited with Grade "A++" by NAAC

JEPPIAAR NAGAR, RAJIV GANDHI SALAI, CHENNAI-600119

APRIL - 2024



SATHYABAMA

INSTITUTE OF SCIENCE AND TECHNOLOGY

(DEEMED TO BE UNIVERSITY)

Accredited "A++" Grade by NAAC | 12B Status by UGC | Approved by AICTE

www.sathyabama.ac.in

DEPARTMENT OF BIOMEDICAL ENGINEERING

BONAFIDE CERTIFICATE

This is to certify that this Project Report is the bonafide work of **MOHAMED AZARUDEEN S (40240019), JEY SURYA S (40240013)** who have done the project work as a team and carried out the project entitled "**LOW COST NON INVASIVE ANALYSIS FOR MEASURING FAT AND WATER CONTENT OF THE BODY**" under our supervision from July 2023 to April 2024.

Internal Guide

Dr. GRACE KANMANI PRINCE M.E., Ph.D

Head of the Department

Dr.T SUDHAKAR M.Sc.,Ph.D.,

Head of the Department
Department of Biomedical Engineering
Sathyabama Institute of Science & Technology
(Deemed to be University)
Category I University by UGC
Jeppiaar Nagar, Rajiv Gandhi Salai,
Chennai - 600 119

Submitted for Viva voce Examination held on 05/04/2024

Internal Examiner

External Examiner

DECLARATION

We **MOHAMED AZARUDEEN S (40240019)** and **JEY SURYA S (40240013)** hereby declare that the Project Report entitled "**LOW COST NON INVASIVE ANALYSIS FOR MEASURING FAT AND WATER CONTENT OF THE BODY**" done by us under the guidance of **Dr. Grace Kanmani Prince M.E., Ph.D** at "**SATHYABAMA INSTITUTE OF SCIENCE AND TECHNOLOGY**" is submitted in partial fulfillment of the requirements for the award of Bachelor of Technology degree in Biomedical Engineering.

DATE: 05/04/2024

PLACE: CHENNAI



1) **MOHAMED AZARUDEEN S**



2) **JEY SURYA S**

SIGNATURE OF THE CANDIDATES

ACKNOWLEDGEMENT

We are pleased to acknowledge our sincere thanks to the **Board of Management of Sathyabama Institute of Science and Technology** for their kind encouragement in doing this project and for completing it successfully. We are grateful to them.

We convey our thanks to **Dr. T. Sudhakar, M.Sc., Ph.D., Head of the Department, Department of Biomedical Engineering** for providing us necessary support and details at the right time during the progressive reviews.

We would like to express our sincere and deep sense of gratitude to our Project Guide **Dr. Grace Kanmani Prince , M.E., Ph.D., Department of Biomedical Engineering** for her valuable guidance, suggestions, and constant encouragement that paved way for the successful completion of our project work.

We wish to express our thanks to all **Teaching and Non-teaching** staff members of the **Department of Biomedical Engineering** who were helpful in many ways for the completion of the project.

We would like to express our thanks to our parents for their unwavering support to us during these years.

ABSTRACT

The project aims to develop a comprehensive body composition measurement system that considers both external parameters like weight and height, as well as internal parameters such as body water and fat content. It starts with creating a non-real-time prototype using ultrasonic sensors for height measurement and load cells for weight measurement. Body mass index should be calculated from height and weight and by using the derived formula the fat percentage could be calculated for external parameters. Additional linear regression analysis is conducted to establish predictive models based on the collected datasets. By integrating technological advancements with established anthropometric measurements, this project aims to contribute to the development of more precise and accessible methods for assessing body fat composition, after establishing precise measurements, the project will transition to real-time monitoring to assess body composition more accurately. The collected data will be analyzed to calculate body fat and water content. Ethical considerations and privacy protections will be paramount throughout the project. Ultimately, the goal is to create a versatile system with applications in healthcare, fitness assessment, and wellness monitoring, providing valuable insights into an individual's health and fitness. Following the establishment of precise measurements, the project transitions to real-time monitoring, incorporating bio impedance analysis to assess body composition more accurately. This includes the analysis of collected data to calculate body fat and water content, further enhancing the system's capabilities in providing comprehensive health assessments.

Keywords: Body composition; Body mass index; fat percentage; Water content.

TABLE OF CONTENTS

CHAPTER NO	CONTENT	PAGE NUMBER
	ABSTRACT	i
	LIST OF ABBREVIATIONS	iv
	LIST OF FIGURES	v
CHAPTER 1	INTRODUCTION	1
CHAPTER 2	LITERATURE SURVEY	4
CHAPTER 3	AIM AND SCOPE OF THE PRESENT INVESTIGATION	20
	3.1 AIM	20
	3.2 SCOPE	20
CHAPTER 4	MATERIALS AND METHODS, ALGORITHM USED	22
	4.1 METHODOLOGY	22
	4.2 HARDWARE REQUIREMENTS	23
	4.2.1 ARDUINO	23
	4.2.2 ULTRASONIC SENSOR	23
	4.2.3 LOAD CELL	24
	4.2.4 POWER SUPPLY	25
	4.2.5 IMAGE PROCESSING MODEL	25
	4.2.6 SYSTEM INTEGRATION	25

CHAPTER NO	CONTENT	PAGE NUMBER
	4.2.7 DATA COLLECTION AND ANALYSIS	25
	4.2.8 USER INTERFACE DESIGN	26
	4.3 METHODS	26
	4.3.1 WEIGHT DETECTION USING LOAD CELL	26
	4.3.2 HEIGHT DETECTION USING ULTRASONIC SENSOR	28
	4.3.3 BODY FAT CONTENT CALCULATION	29
	4.3.4 BODY WATER CONTENT CALCULATION	30
	4.3.5 LINEAR REGRESSION METHOD	31
	4.3.6 DATASET WITH LINEAR REGRESSION	32
	4.4 ALGORITHMS USED	33
	4.4.1 BODY COMPOSITION ANALYSIS	33
CHAPTER 5	RESULTS AND DISCUSSION, PERFORMANCE ANALYSIS	36
	5.1 RESULTS AND DISCUSSION	36
	5.1.1 HEIGHT AND WEIGHT DETECTION WITH ARDUINO	36
	5.1.2 BFP USING LINEAR REGRESSION	37
	5.1.3 BWP USING LINEAR REGRESSION	38
	5.1.4 GRAPHICAL REPRESENTATION OF BFP	39
	5.1.5 HEIGHT AND WEIGHT IN IMAGE PROCESSING	40

CHAPTER NO	CONTENT	PAGE NUMBER
	5.1.6 WEIGHT DETECTION USING IMAGE PROCESSING	41
	5.2 DISCUSSION	42
	5.3 PERFORMANCE ANALYSIS	42
CHAPTER 6	SUMMARY AND CONCLUSION	42
	6.1 SUMMARY	42
	6.2 CONCLUSION	43
	REFERENCES	45

LIST OF ABBREVIATIONS

ABBREVIATIONS	DEFINITIONS
BF	Body Fat
ID	Intellectual disability
DXA	Dual Energy X-Ray
BMI	Body Mass Index
TBM	Total Body Water
BIS	Bioelectrical Impedance Spectroscopy
FFM	Fat Free Mass
ADP	Adenosin Di Phosphate
BIA	Bioelectrical Impedance analysis
LMF	Local muscle fatigue
HBC	Human body composition
MCU	Microcontroller Unit
MRI	Magnetic Resonance Imaging
CPD	Conditional probability
BIVA	Bioelectric Impedance vector
BMC	Bone mineral content
CVD	Cardiovascular Diseases

LIST OF FIGURES

FIGURE NO.	FIGURE NAME	PAGE NO
4.1	BLOCK DIAGRAM OF BFP DEVICE	22
4.2	ARDUINO BOARD	23
4.3	ULTRASONIC SENSOR	24
4.4	LOAD CELL	24
4.5	LOAD CELL CIRCUIT	27
4.6	CIRCUIT FOR HEIGHT AND WEIGHT MEASUREMENT	29
5.1	HEIGHT AND WEIGHT DETECTION RESULT	36
5.2	BFP WITH LINEAR REGRESSION RESULT	37
5.3	BWP WITH LINEAR REGRESSION RESULT	38
5.4	GRAPHICAL REPRESENTATION	39
5.5	HEIGHT DETECTION RESULT	40
5.6	WEIGHT DETECTION RESULT	41

CHAPTER – 1

INTRODUCTION

A comprehensive understanding of an individual's health and wellness is deeply rooted in body composition analysis, a process that delves into the distribution of various components within the body, including fat, water, muscle, and bone mass. These insights gained from body composition analysis are invaluable for healthcare, as they aid in the analysis of an individual's physical condition, inform fitness strategies, and enable personalized wellness tracking. However, while the importance of body composition analysis is undeniable, traditional methods often present significant challenges. Many of these methods demand specialized equipment that can be costly, require trained personnel for accurate measurements, or involve intricate and time-consuming procedures. These barriers have limited the widespread adoption of body composition assessment and have underscored the need for innovative, accessible solutions. This project seeks to address these challenges by pioneering the development of a novel and accessible body composition measurement system. Unlike conventional approaches, this system relies on readily available sensors, potentially revolutionizing the landscape of health and fitness monitoring with non-invasive method. Our motivation to embark on this endeavour is rooted in the recognition that while previous projects have made progress in this field, they have typically focused on singular aspects, such as height and weight, or have employed advanced but complex technologies like Dual-Energy X-ray Absorptiometry (DEXA).

These methods, while accurate, often fall short in terms of practicality for everyday use due to their cost, equipment requirements, and the expertise needed to operate them. The most common, easy-to-use, and cost-effective methods to assess overweight and obesity are based on anthropometric measures such as the body mass index (BMI): weight divided by height squared, and skinfolds. Scientific evidence on the misclassification of obesity using BMI exists, especially in children. BMI does not distinguish between lean muscle from fat mass, nor does it accurately reflect visceral

fat accumulation. Visceral fat is the likely culprit leading to most of the metabolic and clinical consequences of obesity. Ideally, obesity should be determined based on the percent of body fat (%BF), which is considered a better marker for assessing excess adiposity or obesity (Int. J. Environ. Res. Public Health 2022). Our project is designed to build upon the successes and lessons learned from these prior efforts. We emphasize the foundational importance of height and weight as parameters in body composition assessment. We detect the body fat percentage (BFP) by both internal and external parameters.

Our primary objective is to create a system that is both affordable and user-friendly. This innovative approach leverages recent advancements in sensor technology, data processing techniques, and sophisticated analysis algorithms to make body composition assessment accessible to a broader audience. The vision for our project is ambitious but purposeful. We aspire to develop a comprehensive body composition measurement system that empowers individuals to take charge of their health and fitness by providing fundamental body composition data. To achieve this, we will combine the capabilities of an ultrasonic sensor for height measurement with load cells for precise weight measurement. However, we do not intend to stop there. Our system will evolve to enable real-time monitoring, which will allow continuous tracking of changes in body composition. This feature is particularly significant for individuals aiming to monitor their progress in real time and make timely adjustments to their health and fitness routines. Accessibility and privacy are fundamental considerations in our project. In this study, we explore the integration of image processing models and sensor technologies to devise a comprehensive method for measuring body fat composition. We are deeply aware of the sensitive nature of personal health data, and as such, ethical and privacy safeguards will be seamlessly integrated into our system design.

A key aspect of our methodology involves the utilization of image processing techniques to predict an individual's height. By analyzing facial features such as jawline length, eyebrow length, nose length, eye length, outer mouth length, and inner mouth length, we develop a model capable of estimating height with a high degree of

accuracy. We prioritize user consent and data protection to ensure that the information collected is handled responsibly and in line with established privacy standards. In summary, our project represents a promising step towards making body composition assessment accessible, cost-effective, and user-centric. We believe that the successful implementation of our body composition measurement system will not only empower individuals seeking to improve their health and wellness but also find valuable applications in healthcare, sports, and the fitness industry, potentially influencing preventive healthcare measures and the understanding of population health trends.

The significance of this study lies in its potential to revolutionize the way body fat composition is assessed. By leveraging advancements in image processing and sensor technologies, we aim to develop a method that is not only accurate but also accessible and convenient for individuals seeking to monitor their health and fitness levels. Moreover, the integration of these techniques opens possibilities for automation and scalability, facilitating widespread adoption in diverse settings. Through this research, we aspire to contribute to the advancement of personalized healthcare and wellness practices.

CHAPTER - 2

LITERATURE SURVEY

Fabio Bertapelli et al., (2021) proposed using Body Mass Index (BMI), age, and sex as predictors, the set out to create and cross-validate an equation to estimate body fat percentage (BF%) in children with intellectual disability (ID). Additionally, they evaluated the effectiveness of the Deuenberg equation, an established BMI-based calculation for BF% in kids with ID. The researchers discovered a significant link between the Dual-Energy X-Ray Absorptiometry (DEXA)-determined and the BMI-predicted BF%, with similar correlation coefficients and few differences between the two techniques. This indicates that major determinants of %BF in children with ID are BMI, age, and sex.

Malia NM Blue et al., (2021) proposed examining the body composition among individuals from different racial and ethnic backgrounds, variations in the mentioned components have been observed across different groups. This analysis's main objective was to evaluate the precision of widely-used body composition assessment tools in samples made up of people of different races and ethnicities, as well as in African-American, Hispanic, Asian, and Native American populations. Existing literature indicates that air displacement plethysmography (ADP) is valid for Hispanic and African American males when using race-specific equations, whereas dual-energy X-ray absorptiometry (DEXA) yields reliable results when utilized in multi-ethnic populations. Additionally, research shows that results between races and ethnicities can vary. When applying race-specific equations, bioelectrical impedance analysis (BIA) may produce valid results in Hispanic, Native American, and Asian populations. However, it is important to note that for African-American/Black individuals, current data does not support the recommendation of BIA due to the lack of validation through a comprehensive model criterion. This validation process is crucial in order to better comprehend the risk of diseases within society as a whole, enhance exercise and dietary recommendations for disease prevention and management, and effectively track changes resulting from lifestyle interventions.

Dongming Li et al., (2020) proposed Electro-Impedance Myography (EIM) electrode configuration was introduced and validated for evaluating lower limb muscle fatigue (LMF). A four-layer human upper arm simulation model in Finite Element Method (FEM) software was established to optimize muscle selectivity. Sensitivity parameters were introduced to reduce the impact of individual variations in fat thickness on EIM results. Dynamic and static muscle fatigue experiments were conducted on biceps brachii muscles in 10 volunteers to validate the proposed electrode configuration. The new configuration reduced the measurement area by 25% while maintaining impedance amplitude and sensitivity. It significantly mitigated the impact of fat thickness differences on EIM results. The study found that as fatigue increased, resistance decreased, and this change was more pronounced with heavier loads. These findings offer valuable insights into the development of wearable devices for assessing LMF, providing both technical support and theoretical guidance.

Maria F. Bernal-Orozco et al., (2020) proposed analyze the anthropometric and body composition profiles of young professional soccer players, making comparisons across different competitive divisions and playing positions. The research involved measuring body mass, height, sitting height, skinfolds, girths, and bone breadths, with body composition assessed using a 5-component model. Results from 755 subjects were examined. Notably, the U-20 division exhibited higher values compared to other divisions, while the fourth division, representing the youngest category, displayed lower measurements for most body composition metrics and average anthropometric variables. Goalkeepers, on the other hand, tended to have greater body mass, height, sitting height, skinfold sum, and body adipose mass, both in terms of absolute values and percentages. It's important to note that the methodology for determining body composition used a 5-component model, making distinctions from traditional bi-compartmental methods.

Dean Henry S. Guerrero et al., (2020) proposed the device used in this investigation is built with the goal of producing signals that will go through the human body. Additionally, it has an amplifier to boost the 10-microampere current signal, which is relatively low for safety reasons. The goal of the project is to develop a nondestructive method of calculating a human's body fat percentage. The goal of the project is to develop a nondestructive tool that might measure a person's body fat percentage. The constructed gadget was able to measure the necessary body fat ranges using a non-invasive electrode and a Gizduino 328 microcontroller-unit with results that were essentially identical to those of the skin caliper approach. The result of the created BIA device was acceptable based on the comparative testing against the conventional method which is the skin caliper method in measuring the body fat percentage since it only yields a 10% percentage error.

Yao Lu et al., (2019) proposed the supervised inference model to predict pixel-level body composition and percentage of body fat using 3D geometry features and body density they used body density to model a fat distribution base prediction. The probability of the base prediction bias is then inferred with the help of 3D geometry information using a Bayesian network. Finally, they use non-parametric regression to correct the bias. When building and validating their models, they leverage DXA assessment as the source of truth. On average, their approach outperforms those prediction models by 52.69%. In terms of determining the whole-body fat percentage, they contrast our method with BOD POD, a piece of medical equipment. By 23.28%, our approach outperforms the BOD POD. The moderate sample size is a significant flaw in their research.

Lynn Moeng-Mahlangu et.al., (2022) proposed by comparing %BF determined by bioelectrical impedance analysis (BIA) and calculated from available prediction equations based on skinfolds in young South African children. A cross-sectional study performed on 202 children (83 boys and 119 girls) aged 6–8 years. All three skinfolds' equations demonstrated a smaller %BF than the %BF measured by BIA at it was able to compare %BF derived from secondary methods of skinfold thickness with BIA even

though it does not include 'gold-standard methods in a relatively large and homogenous sample of young African children. However, in the absence of developed %BF equations or 'gold-standard' methods, the available predictions are still desirable.

Bassem Ibrahim et.al (2021) proposed, a Bio-Z simulation platform is presented that uses a 3D grid of time-varying impedance voxels to simulate tissue and artery pulse wave activity. It simulates voltage distribution at nodes and current flow through grid elements, including tissue DC voltage and blood flow pulse signals. The evaluation of bio-impedance signal penetration at different artery depths is made possible by validation, which takes into account electrode placement, current frequency, and size impacts on detected pulse signals. This simulation tool offers direction for circuit designers and algorithm developers, improves Bio-Z sensing for measuring hemodynamic parameters, and helps us understand how pulse waves propagate through tissues.

Myoung Hoon Jung et al., (2021) proposed a wrist-wearable bioelectrical impedance analyzer with a contact resistance compensation function such that bioelectrical impedance can be accurately estimated even with considerably small sizes of electrodes (outer electrodes: 68 mm²; inner electrodes: 128 mm²). The correlation coefficient and the SEE of percentage body fat relative to the DEXA instrument were estimated to be 0.899 and 3.8%BF, respectively, which are above the level of performance of the commercial upper-body portable body fat analyzer. Considering that the measurement time of our wrist-wearable BIA device was only 7 s and could be reduced further, this sensor technology provides a new possibility for a wearable bioelectrical impedance analyzer with more miniature electrodes toward daily obesity management.

Sergey G. Rudnev and Elena Z. Godina (2022) proposed body composition assessment primarily relies on anthropometry and bioimpedance analysis, with limited use of methods like DXA, primarily in clinical research. The Matiegka four-component tissue-system model is still widely utilized despite its incompatibility with molecular-level

models. The country's body composition research has faced challenges due to a missed technological advancement during the late 1980s, and there's a lack of validation studies using reference methods, resulting in the use of foreign regression formulas for indirect methods. Notably, advanced techniques like the deuterium dilution method and 3D laser-based scanning are being explored, alongside mass population bioimpedance measurements in health centers. While Russia has made progress, the absence of conventional reference methods hinders international recognition.

Erika Aparecida Silveira et.al.,(2020) proposed that when compared to our standard reference, the DXA, the BIA and the SF equation exhibited a significant level of concordance to estimate body fat percentage in all individuals and among women. However, in both men and women with high body fat percentages, both approaches underestimated their body fat proportion. DXA and the anthropometric equation created by Durnin and Womersley showed a strong level of concordance in men, but BIA exhibited a moderate level of concordance in this group. Future studies on older persons should take into account a variety of techniques, varied ethnic groups, and the creation of fresh SF equations.

Linauskas et.al (2019) proposed a study involving 210 men and 456 women over a 20.1-year follow-up period revealed varying associations between body composition and rheumatoid arthritis (RA) risk. In women, a 10% higher RA risk was linked to each 5% increase in total body fat, a 5% higher risk for every 5cm increase in waist circumference (WC), and nearly a 50% higher risk for those with an obese BMI compared to normal. These associations extended to 'other RA.' However, in men, no clear connections were found between body fat percentage, WC, or BMI and RA. Notably, no significant associations were established for 'seropositive RA' in either women or men, potentially due to limited sample size.

Seung-Chul Shin et.al., (2019) proposed a faster and more accurate method that utilizes a small dry electrode-based wearable device, which predicts whole-body impedance using only upper-body impedance values. Such a small electrode-based

device typically needs a long measurement time due to increased parasitic resistance, and its accuracy varies by measurement posture. They created a sensing mechanism that only makes touch with the wrist and index fingers in order to reduce these variations. An efficient parameter calibration network also decreased the measurement time to five seconds. Then, using lower-body anthropometric data and measurements of upper-body impedance as auxiliary input features, they constructed a deep neural network-based algorithm to predict the PBF value. The findings revealed a 9% improvement in the correlation coefficient (r^2) value and a 28% decrease in the standard error of estimation (SEE).

Francesco Campa et.al (2021) proposed on body composition analysis, particularly emphasizing bioelectrical impedance analysis (BIA) and bioelectrical impedance vector analysis (BIVA) in athletes. It discusses the use of specific technologies and sampling frequencies, offering recommendations for assessing body composition in athletes. The review explores both quantitative parameters (e.g., fat mass, body fluids) and qualitative parameters (e.g., vector position, phase angle) for tracking changes in body composition during different training cycles, aligned with the competitive season. Notably, assessments should be conducted after a consistent time lag from training or competition, providing valuable insights into athletes' physical adaptations and performance optimization.

Erika Severeyn et.al (2020) proposed the evaluation of the capability of the anthropometrics variables to discriminate subjects with abnormal BFP. The study made use of a database containing 1053 participants and 28 anthropometric measurements. To assess the predictive power of the anthropometric factors evaluated, the area under the receiver operating characteristic curves (AUCROC), sensitivity (SEN), specificity (SPE), and negative predictive value (NPV) were determined. The variables with the best ability to detect aberrant BFPs were four skinfolds (calf, suprailiac, abdomen, and thigh) and three circumferences (arm, waist, and hip), with an $AUCROC > 0.800$ ($SEN > 0.760$ and $SPE > 0.673$). Because of this, abdominal and suprailiac fold

anthropometric measurements had a high likelihood of identifying normal-weight patients.

Katherine Sullivan et.al (2020) proposed to compare body volume (BV) estimated from a 2-dimensional (2D) image analysis program (BVIMAGE), and a dual-energy x-ray absorptiometry (DXA) equation (BVDXA-Smith-Ryan) to an underwater weighing (UWW) criterion (BVUWW) and to compare relative adiposity (%Fat) derived from a 3-compartment (3C) model using BVIMAGE (%Fat3C-IMAGE), and a 4-compartment (4C) model using BVDXA-Smith-Ryan (%Fat4C-DXA-Smith-Ryan) to a 4C criterion model using BVUWW (%Fat4C-UWW). Forty-eight participants were included (60% male, 22.9 ± 5.0 years, 24.2 ± 2.6 kg/m²). BVIMAGE was derived using a single digital image of each participant taken from the rear/posterior view. The 2D image analysis program provided an accurate and non-invasive estimate of BV and subsequently %Fat within a 3C model in generally healthy, young adults.

N.I. Khramtsova and S.A. Plaksin (2020) proposed the diagnostic value of bioelectrical impedance analysis (BIA) was explored to assess changes in human body components before and after water-jet liposuction in 37 healthy women. Using a portable non-professional analyzer, "Tanita BC-542," measurements were taken for subcutaneous adipose tissue, body water, muscle, bone components, and basal metabolism degree. Notably, body weight initially increased significantly post-liposuction but returned to baseline by the 7th day, linked to surgical fluid infusion. The study revealed correlations between changes in body fat percentage, the volume of removed adipose tissue, and BIA measurements immediately after surgery and on the 7th day, highlighting the utility of BIA in tracking dynamic body composition changes.

César Leão et al.,(2019) proposed a comprehensive analysis of the anthropometric profiles of a sizable cohort of soccer players, categorized by both age groups (12-14, 14-16, 16-18, and 18-37 years) and playing positions. The study evaluated weight, height, and skinfold thickness measurements. For players aged 18 and older, Reilly and Evans formulas were employed to estimate body fat percentages. The research

revealed that anthropometric characteristics varied significantly depending on the players' positions across various age groups. Notably, the study also highlighted the substantial impact of different validated equations on the estimated body composition, underscoring the complexity of assessing body composition in adult soccer players.

Javier Alves et al., (2021) proposed Several anthropometric and performance parameters related to aerobic metabolism are associated with performance in endurance runners and are modified according to the training performed. The objective of this study was to investigate the ergospirometric and body composition changes in endurance runners during a sports season in relation to their training. Twenty highly trained men endurance runners performed an incremental test until exhaustion (initial, and at 3, 6, and 9 months) on a treadmill to determine maximal oxygen consumption second ventilatory threshold, and their associated running speeds. Skinfolds, perimeters, and weights were measured. No changes were obtained in or during the study, although their associated running speeds increased ($p < 0.05$) after 3 months of the study. Decreases in fat mass ($p < 0.05$) and muscle mass ($p < 0.05$) were observed at the end of the season (9 months). Changes occurred in the different skinfolds according to the characteristics of the training performed during the season. In conclusion increase with a greater volume of kilometres trained and can be adversely affected by loss of muscle mass.

Srinivasa Bhattachar et al., (2021) proposed the relationship between body fat content and maximum aerobic capacity was investigated in 104 healthy Indian males. Various methods, including body mass index (BMI), bioimpedance, skinfold thickness (SFT), body girth (BG) measurements, waist circumference (WC), and waist-hip ratio (WHR), were used to assess body fat content. Maximum aerobic capacity (VO₂max) was determined indirectly from maximum heart rate during a treadmill test. The study found that when expressed in liters per minute (L/min), VO₂max exhibited a positive correlation with body fat content regardless of the estimation method. However, when expressed in milliliters per kilogram per minute (ml/kg/min), VO₂max showed a negative correlation with five out of seven clinical fat estimation parameters.

Jihan K. Boughman et al., (2019) proposed twenty-two NOLS participants completed a 23-d backpacking trip into the Wind River Range in Wyoming. Pre- and postexpedition anthropometric measures were collected using 3 different body composition measurement tools: Tanita segmental body composition monitor, Harpenden skinfold calipers, and the Bod Pod. For the purposes of this study, the Bod Pod was used as the standard against which other methods were compared. This study shows that the skinfold caliper and the Tanita scale give measurements comparable to the less readily available and more costly Bod Pod measurement in backcountry expedition participants. Thus, the use of these tools can provide outdoor programs with valid measurements that can be used as a reference point to help ensure participants are adequately meeting energy requirements.

Hyeuknam Kwon et.al (2019) proposed the Electrical impedance myography (EIM) is a relatively new technique to assess neuromuscular disorders (NMD). Although the application of EIM using surface electrodes and has been adopted by the neurology community in recent years to evaluate NMD status, sensitivity as a biomarker of skeletal muscle condition is impacted by subcutaneous fat (SF) tissue. Here, we develop a method that is able to remove the contribution of SF from sEIM data. evaluate independent component analysis (ICA) and principal component analysis (PCA) for this purpose. Then, we introduce the so-called model component analysis (MCA). All methods are validated with numerical simulations using impedivity data from SF and muscle tissues. The methods are then tested with measurements performed in diseased individuals. Whereas the Simulations demonstrate that MCA is the most accurate method at separating the impedivity of SF and muscle tissues with the accuracy being 99.2%, followed by ICA, finally PCA with Experimental results from sEIM data measured on the triceps brachii of patients are consistent with muscle grayscale level values obtained using ultrasound imaging controls. MCA can be used to separate the impedivity of SF and muscle tissues from sEIM data, thus increasing the sensitivity to detect changes in the muscle..

F.Javier Nunez et al., (2020) proposed to determine the most effective field method to estimate FFM changes in elite young soccer players. similar predictors of DXA-derived FFM changes between pre- and mid-season. To the author's knowledge, no previous investigation has analyzed these field methods to estimate FFM changes throughout the season in elite youth soccer players. A previous study with rugby league players showed that the lean-mass index was very likely to be more useful than BIA for a better estimate of FFM changes. A recent study with elite youth soccer players also showed that five anthropometric equations were more useful than BIA for estimating of FFM changes. Lean-mass index and BIA (inbody) showed statistical differences in biases in comparison with DXA. the biases indicate that these should be used for estimating FFM changes between pre- and mid-season in youth soccer players. BIA and all anthropometric equations used in this study to estimate FFM change showed high correlations with DXA data

Susanna C. Larsson et al., (2020) proposed Body mass index and body composition in relation to 14 cardiovascular conditions in UK biobank. The Mendelian randomization is an epidemiologic method based on instrumental variable analysis and utilizes genetic variants, usually single nucleotide polymorphisms (SNPs), associated with the modifiable risk factor which is BMI as proxy indicators to infer causality is 12 where the MR design reduces bias due to confounding as alleles are randomly assorted at conception and prevents reverse causation bias because disease cannot affect genotype. In this study they assessed the causal roles of BMI and body composition for a wide range of cardiovascular conditions and found evidence that higher BMI and especially fat mass index are associated with increased risk of most outcomes. A novel finding is a strong and consistent positive association between genetically predicted BMI and aortic valve stenosis. In contrast, there was significant or suggestive evidence of inverse associations of fat-free mass index with atrial fibrillation, ischaemic stroke, and abdominal aortic aneurysm. With regard to body composition the results are genetically predicted fat mass index and atrial fibrillation confirm the results of the fat mass was positively associated with atrial fibrillation.

Shengkui Zhang et al., (2021) proposed the Association of Rotating Night Shift Work with Body Fat Percentage and Fat Mass Index among Female Steelworkers in North China. to evaluate the associations of rotating night shift work with body fat percentage (BF%) and fat mass index (FMI). A cross-sectional study was conducted among 435 female steelworkers, aged 26–57 years in Tangshan, China. BF% was assessed via bioelectrical impedance analysis and FMI was calculated. Different exposure metrics of night shift work were used to examine the effects of night shift work on BF% and FMI. The duration (years), cumulative number (nights), and cumulative length of night shifts (hours) were positively correlated with FMI and BF%, and these relationships were independent of body mass index (BMI). Compared with day workers, night shift workers with an average frequency of night shifts and Non obese night shift workers by the BMI criterion should also be alert to the risk of the excess accumulation of body fat, which is actually responsible for most obesity-associated adverse health consequences. Night shift workers with a relatively “normal weight” should still be alert to the risk of the excess accumulation of body fat, which is actually responsible for most obesity-associated adverse health consequences. Health interventions for related populations need to be improved, which is currently more focused on overall weight control.

Ningjian Wang et al., (2021) proposed the Total and regional fat-to-muscle mass ratio measured by bioelectrical impedance and risk of incident type 2 diabetes total and regional FMRs were independently associated with the risk of type 2 diabetes also it included a large sample size and relatively long follow-up duration, which enabled adequate power to study the associations and interactions in detail. We collected various covariates, including demographic information, lifestyle factors, metabolic factors, and medications, which allowed for relatively rigorous adjustments. FMRs of the whole body and leg showed the strongest associations with incident diabetes in men and women, respectively. These associations were obviously strengthened among participants with BMI over 25 kg/m². Some factors affecting bioimpedance results, such as hydration, room temperature, or prior exercise, were not intentionally controlled. Because it normally required more than 1 hour to complete the assessments

before the bioimpedance measurement in the UKB assessment centre with stable temperature, most of the participants were likely to be in a steady state when reaching the bioimpedance room. The clinical implications might be to stratify strategies for diabetes prevention by targeting lowering fat mass and increasing muscle mass, especially among people with BMI over 25 kg/m².

Antonios h. tzamaloukas et al.,(2003) proposed the estimation of body water, fat-free mass, and body fat in patients on peritoneal dialysis by anthropometric formulas by comparing the estimation of body water, fat-free mass, and body fat by anthropometric formulas to estimates employing dilution of tracer substances to measure body water and standard methods to analyse body composition in studies performed on peritoneal dialysis patients they also analysed the potential errors of the estimates of body composition by the formulas therefore Estimates of the average body composition provided by the anthropometric formulas agreed with estimates provided by the standard methods. However, these formulas have the potential of introducing large errors when estimating body composition in individuals differing from the average subject composition in individuals differing from the average subjects either because the anthropometric formulas do not account for major determinants of body composition, such as physical exercise, nutrition, and catabolic illness, or because these formulas systematically overestimate body water in subjects who are obese or experiencing volume excess. The potential errors include the estimates of body water. These errors may alter the interpretation of urea kinetic studies in certain categories of peritoneal dialysis patients such as obese subjects.

Seda Ugras (2020) proposed the altered hydration status on effectiveness of body composition analysis using bioelectric impedance analysis by providing valuable information about the assessment of balance between fluid, fat, and fat-free mass. The present study assessed whether the regular increase in water consumption has any significant effects on measurements of body composition using BIA. A total of 140 subjects participated in this study. Body composition analyses of the subjects were

performed for five times using foot to foot BIA first at baseline and four times, after 15 minutes of each 500 mL water intakes. There was overestimation in body fat mass compared to the baseline and increased significantly after water intakes in males and in females. In contrast, there was markedly underestimation of total body water and fat-free mass in males (35% and 50%) and females (17% and 22%), respectively at the end of the study thus the overestimation of body fat mass and lower estimation of total body water and fat-free mass with closely to the amount of water retention using BIA application should be considered in health-related clinical practice. Acute water application caused significantly alteration in body composition analyses using BIA method. These variations in body composition compartment also varied between male and female subjects, and increased water intake causes more increase in %FM and FM in females than those of male subjects.

Willem G.P.M. Looijaard et al., (2019) proposed the low muscle mass identifying in critically ill patients based on Agreement between bioelectrical impedance analysis and computed tomography. During the study period 1654 patients were admitted to the ICU with a mean age of 63 ± 16 years and a mean APACHE IV predicted mortality of $24.5 \pm 30.1\%$. One hundred eighty-five patients fulfilled the inclusion criteria. In nine patients, the BIA measurement could not be performed within 72 h of the CT scan. Forty-five patients were excluded because their CT-scan was unsuitable for analysis due to artefacts, scattering, muscles cut-off due to windowing, or insufficient scan quality; 12 patients because BIA could not be performed due to the presence of internal or external metal devices and 9 patients due to erroneous BIA values because of shivering, dyspnea, or the inability to obtain correct positioning. A total of 110 patients were included in final analyses. The BIA measurement was performed a median of 1 day after CT (0–1), and in 97 patients within one day of the CT scan (88%). Sensitivity analyses were performed in patients in whom the BIA measurement was performed within one day of the CT scan and in those with a reliable weight.

Hanneke Pierre Franciscus Xaverius Moonen et al.,(2020) proposed the association of bioelectric impedance analysis body composition and disease severity in COVID-19 hospital ward and ICU patients were performed an observational cross sectional cohort study. Bioelectric impedance analysis was conducted amongst all confirmed COVID-19 patients admitted to the ward or ICU of our hospital in the Netherlands, between April 10 and 17, 2020. Body water measurements and derived values were recalculated to dry weight, using a standard ratio of extracellular water to total body water of 0.38. Data were compared between the ward and ICU patients, and regression models were used to assess the associations between baseline characteristics, body composition, and several indicators of disease severity, including a composite score composed of mortality, morbidity, and ICU admission. Fifty-four patients were included, of which thirty in the ward and twenty four in the ICU. Thus, the Body composition values were not independently associated with disease severity. In multiple logistic regression analyses, a low phase angle was associated with COVID-19 severity in the composite score. Therefore they found no significant associations between body composition, including fat mass, visceral fat area, and fat-free mass, and disease severity in our population of generally overweight COVID-19 patients. A lower phase angle did increase the odds of severe COVID-19. We believe that factors other than body composition play a more critical role in the development of severe COVID-19.

Unhee Lim et al.,(2018) proposed the propensity for Intra-abdominal and Hepatic Adiposity Varies Among Ethnic Groups by comparing the fat storage in the abdominal region among individuals from 5 different ethnic/racial groups to determine whether fat storage is associated with disparities observed in the metabolic syndrome and other obesity-associated diseases.so they collected data from 1794 participants in the Multiethnic Cohort Study (60–77) years old of African, European (white), Japanese, Latino, or Native Hawaiian ancestry) with body mass index values of 17.1–46.2 kg/m².From May to April, participants visited the study clinic to undergo body measurements, an interview and a blood collection. Participants were evaluated by dual energy X-ray absorptiometry and abdominal magnetic resonance imaging. Among

the ethnic groups, compared adiposity of trunk, intra-abdominal visceral cavity, and liver, adjusting for total fat mass and evaluated the association of adult weight change with abdominal adiposity; and we examined the prevalence of the metabolic syndrome mediated by abdominal adiposity. In an analysis of data from the participants in the Multiethnic Cohort Study they found extensive differences among ethnic racial groups in the propensity to store fat intra-abdominally. This observation should be considered by clinicians in the prevention and early detection of metabolic disorders.

Zongwen Fan (2022) proposed Obesity associated with having excess body fat, is a critical public health problem that can cause serious diseases. Although a range of techniques for body fat estimation have been developed to assess obesity, these typically involve high-cost tests requiring special equipment. Thus, the accurate prediction of body fat percentage based on easily accessed body measurements is important for assessing obesity and its related diseases. By considering the characteristics of different features (e.g. body measurements), this study investigates the effectiveness of feature extraction for body fat prediction.

Matheus Pachec et al., (2021) proposed the relation between body mass index and body fat percentage in Brazilian adolescents: assessment of variability, linearity, and categorization. Reliable but accessible measures to capture body composition are highly important as the world is in an era of obesity-increase. The most used measure, the body mass index (BMI), nevertheless, has been judged as non-reliable to estimate body fat percentage (BF%). The present manuscript assessed the criticisms of BMI as a predictor of BF% and the BMI-based categories of nutritional status. 4164 children/adolescents from 9 to 18 years of age from Porto Velho, Brazil, had their anthropometric and skinfold measures taken. Controlling for socioeconomic status, school (private, public), sex and age, they compared proposed models/variables in the literature relating BMI and BF%. they evaluated the functions and the residual data to understand the variability of BF% estimate per BMI and evaluated three possible categorization from BMI to predict BF% nutritional status. The function utilizing (linear) BMI was the best to predict BF% ($R^2 = 0.70$) with a variability of only 6.49% around the function. Nevertheless, no categorization of nutritional status was reliable to predict the

nutritional status of individuals. BMI is reliable to estimate BF%. Nevertheless, new normative values must be proposed; the original categorization fails to capture the nutritional status of children/adolescents from this region.

Hyejung Choi (2021) proposed the Body fat-related differences in gait parameters and physical fitness level in weight-matched male adults. Increased body weight is associated with an increased magnitude of foot pressures. Although body mass index (BMI) has been widely used in the assessment of obesity, BMI does not differentiate between muscle and adipose tissue, which may play an important role in characterizing walking patterns. The purpose of the study was to compare gait biomechanics between obese and weight-matched control group

CHAPTER - 3

AIM AND SCOPE OF THE PRESENT INVESTIGATION

3.1 AIM:

The primary aim of this project is to develop an innovative and comprehensive system for measuring body fat composition using advanced sensor technologies and computer vision techniques. To measure fat mass, fat free mass and water content of the body. By integrating ultrasonic sensors, load cells, and image processing models, the project seeks to provide accurate and convenient methods for assessing key biometric parameters including height, weight, and body fat composition.

3.2 SCOPE:

1. **Body Fat Composition Analysis:** The project focuses on developing algorithms and methodologies for accurately measuring body fat composition using a combination of height and weight measurements. By leveraging ultrasonic sensors and load cells, the system aims to provide reliable estimates of body fat percentage tailored for both adult men and women.

2. **Height Measurement using Computer Vision:** Utilizing image processing techniques, the system aims to predict an individual's height based on facial features extracted from images or video frames. By analysing specific facial landmarks, such as jawline length and eye distance, the system provides a non-invasive and automated method for height estimation.

3. **Weight Detection using Load Cell:** Incorporating load cell technology, the project enables precise measurement of weight in real-time. By interfacing with load cells, the system provides accurate weight readings, essential for comprehensive body composition analysis and monitoring.

4. Integration of Sensor Technologies: The project integrates ultrasonic sensors, load cells, and image processing models into a cohesive system for holistic body composition analysis. By combining data from multiple sensors, the system offers a comprehensive assessment of biometric parameters, including height, weight and body fat composition.

5. User Interface and Feedback :The system includes a user-friendly interface, such as an LCD display, to provide real-time feedback on biometric measurements. Users can conveniently monitor their height, weight, and body fat composition, facilitating informed decision-making regarding health and fitness goals.

6. Potential Applications: The developed system has potential applications in various domains, including healthcare, fitness, and wellness management. It can be utilized in clinics, gyms, and home settings for personalized health assessment and monitoring.

7. Scalability and Future Development: The project lays the foundation for further research and development in the field of body composition analysis. Future enhancements may include additional sensor integration, machine learning algorithms for improved accuracy, and compatibility with wearable devices for continuous monitoring

CHAPTER - 4

MATERIALS AND METHODS USED

4.1 METHODOLOGY

The process involves connecting the load cell and ultrasonic sensor to an Arduino microcontroller for persons height and weight (biometric) measurement. The load cell is connected to the Arduino using analog pins, with the HX711_ADC library utilized to communicate with the load cell and retrieve weight measurements. Calibration procedures are conducted to establish the calibration factor for accurate weight readings. Meanwhile, the ultrasonic sensor is connected to the Arduino using digital pins, with the HCSR04 library facilitating distance measurements. Calibration is performed to establish the correlation between distance readings and actual height. In Fig 4.1 shows the block diagram, Arduino continuously samples data from the load cell and ultrasonic sensor, obtaining weight and height measurements, respectively. These measurements are processed using algorithms implemented on the Arduino, which calculate body mass index (BMI) from height and weight and estimate body fat percentage based on predetermined formulas. The resulting biometric data are then displayed on an LCD screen for real-time monitoring and analysis.




FIG: 4.1: Block Diagram of BFP device

4.2 HARDWARE REQUIREMENTS:

- Arduino
- Ultrasonic Sensor
- Load Cell
- Power Supply

4.2.1 ARDUINO UNO:

We using Arduino Uno board here that serve as a central control unit. It serves as a critical component of our BIA system, responsible for data collection, processing and communication. While interfacing the ultrasonic sensor and load cell with Arduino we can collect the data (height and weight) of a subject.




FIG. 4.2 Arduino Board

4.2.2 ULTRASONIC SENSOR:

- It operates by emitting a burst of high frequency sound waves (ultrasonic pulses) and has a receiver to capture the echoes.
- In our project we primarily used ultrasonic sensor to measure the height of the subject by (Total length - distance of an object.)

- It measures the distance from the top of the subjects head to the bottom of their feet.
- The sensor collects the data by recording the time delay between emitting a pulse and receiving it's echo.
- This time delay is used to calculate the distance between the sensor and the object.



FIG. 4.3 Ultra Sonic Senor

4.2.3 LOAD CELL:

- We use load cell to measure the weight of the subjects.
- It does so by detecting the strain or deformation caused by the subjects weight.
- The load cell generates a small electrical signal(mV) proportional to the applied weight.
- The output is acquired by the data acquisition system.



FIG.4.4 Load Cell

4.2.4 POWER SUPPLY:

The use of a battery as a power supply in a Body Fat Analyzer using Arduino offers several key advantages, making it a suitable choice. Lithium batteries are known for their high energy density, which allows for a compact and lightweight power source.

4.2.5 IMAGE PROCESSING MODEL:

Open CV Library (Python) is used to detect the height of a person and weight by their Facial landmarks are detected using the Haar Cascade classifier. Image processing techniques are applied to extract specific facial features such as jawline length, eyebrow length, and eye distance. A custom model is developed to predict height based on these features.

4.2.6 SYSTEM INTEGRATION:

Hardware Used are- Arduino Board, LCD Display.

Software Used here- Arduino IDE.

Ultrasonic sensor, load cell, and image processing model are integrated into a cohesive system. Arduino acts as the central controller, collecting data from sensors and processing it to estimate height, weight, and body fat composition. The LCD display provides real-time feedback to the user.

4.2.7 DATA COLLECTION AND ANALYSIS:

Data Collection: Measurements of height, weight, and body fat composition are collected using the integrated system. Data is recorded for a diverse range of individuals to capture variations in body composition.

Data Analysis: Statistical analysis is performed to evaluate the correlation between sensor readings and actual biometric parameters. Regression analysis may be conducted to refine prediction models and enhance accuracy.

4.2.8 USER INTERFACE DESIGN:

Where the lcd display used as a user interface is designed to display real-time biometric measurements, including height, weight, and body fat composition. Clear and concise feedback is provided to the user for easy interpretation.

4.3 METHODS:

4.3.1 Weight Detection using Load Cell :

we implement the weight detection using a load cell which shown in Fig.4.5 where a transducer that converts force (weight) into an electrical signal. Load cells are commonly employed in various applications requiring precise weight measurements, such as industrial scales, medical devices, and now, in our body composition analysis system.

The provided code utilizes the HX711_ADC library to interface with the load cell. During setup, the Load Cell object is initialized with the appropriate pins connected to the load cell's data (DT) and clock (SCK) pins. Additionally, calibration parameters such as stabilization time and calibration factor are set to ensure accurate weight measurements. Within the loop function, the Load Cell object continuously updates and retrieves eight detection with load cell. retrieved weight value is then displayed on a Liquid Crystal Display (LCD) screen for real-time monitoring. By printing the weight value to the LCD, users can easily visualize and track their weight as part of the body composition analysis process.

This integration of load cell technology enhances the comprehensiveness of our body composition analysis system by providing accurate weight measurements in conjunction with other biometric data such as height. With the ability to monitor weight

changes over time, users can gain valuable insights into their overall health and fitness progression.

Moreover, the non-invasive nature of load cell-based weight detection makes it suitable for integration into wearable devices or home health monitoring systems, further expanding its potential applications in personalized healthcare.




FIG.4.5 LOAD CELL CIRCUIT

4.3.2 Height detection using ultrasonic sensor:

In this section, we employ an ultrasonic sensor to measure the height of an individual. The ultrasonic sensor operates on the principle of sound wave reflection to determine the distance between the sensor and the object being measured. The code provided utilizes the HCSR04 library to interface with the ultrasonic sensor. Initially, the setup function initializes the LCD display and calibrates the sensor by averaging distance readings over a brief period. This calibration ensures accuracy and stability in subsequent height measurements.

Within the loop function, the distance raw variable is updated with the current distance reading from the ultrasonic sensor. By subtracting this value from the initial distance obtained during calibration, the distance object variable represents the height of the object above the sensor. To ensure non-negative height values, a conditional statement is used to set distance object to zero if it falls below zero.

The Fig.4.6 shows the height measurement, represented in centimeters, is displayed on the LCD screen alongside other relevant data such as weight (measured using a load cell) and the raw distance reading from the ultrasonic sensor.

Overall, this approach offers a straightforward and non-contact method for height measurement, suitable for integration into systems requiring real-time monitoring of human biometrics. By combining ultrasonic sensing with other sensor technologies, we aim to develop a comprehensive solution for assessing body composition and health metrics.




Fig 4.6 Circuit for height and weight measurement

4.3.3 BODY FAT CONTENT CALCULATION:

The estimation of a person's body fat percentage typically involves using specific formulas that consider their Body Mass Index (BMI), age, and gender.

These are the formulas for estimating body fat percentage in adult males and females:

For Adult Males:

$$\text{BFP} = 1.20 \times \text{BMI} + 0.23 \times \text{age} - 16.2$$

For Adult Females:

$$\text{BFP} = 1.20 \times \text{BMI} + 0.23 \times \text{age} - 5.4$$

These formulas require the following variables:

- BFP: This represents the Body Fat Percentage, which is what you're trying to estimate.
- BMI: BMI is the Body Mass Index, a measure of how a person's weight relates to their height. It's calculated as weight in kilograms divided by height in meters squared.
- Age: You need to know the person's age in years.

The formulas include specific constants (1.20, 0.23, 16.2 for males and 5.4 for females) to fine-tune the estimation based on the individual's gender.

To estimate body fat percentage begin by calculating their BMI. Then apply the appropriate formula based on their gender, factoring in their age and the constants mentioned. The result will give an approximate body fat percentage.

4.3.4 BODY WATER CONTENT CALCULATION:

The formula of **chumlea et.al** for Caucasian men defines a linear relationship between body weight (W) and body water (V). The estimation of a person's body Water content percentage typically involves using specific formulas that consider their Body Mass Index (BMI) and age.

The formula Used to find water content in a person,

$$V = 23.04 + 0.50 W - 0.62 \text{ BMI} - 0.03 A$$

Where,

V-Body water;

W-Body weight;

A-Age

4.3.5 Linear regression method:

Overview of Linear Regression:

Linear regression is a statistical method used to model the relationship between a dependent variable and one or more independent variables by fitting a linear equation to observed data. It is commonly employed for prediction and forecasting tasks, as well as in the analysis of relationships between variables.

Key Concepts:

1. Dependent Variable (Y): The variable being predicted or explained by the model. In the context of body fat composition analysis, the dependent variable could be body fat percentage.

2. Independent Variable(s) (X): The variable(s) used to predict or explain the dependent variable. In the context of body fat composition analysis, independent variables could include height, weight, age, gender, etc.

3. Linear Equation: The linear regression model assumes that the relationship between the dependent and independent variables can be represented by a straight line equation of the form: $Y = \beta_0 + \beta_1 X_1 + \beta_2 X_2 + \dots + \epsilon$,

Where, Y is the predicted value of the dependent variable,

- β_0 is the intercept (the value of Y when all independent variables are zero),
- $\beta_1, \beta_2, \dots$ are the coefficients (also known as slopes) representing the change in Y for a one-unit change in the corresponding independent variable,
- $X_1, X_2, \dots$ are the values of the independent variables,
- ϵ is the error term representing the difference between the observed and predicted values of Y.

4. Fitting the Model: The goal of linear regression is to estimate the coefficients ($\beta_0, \beta_1, \beta_2, \dots$) that best fit the observed data. This is typically done using the method of least squares, which minimizes the sum of the squared differences between the observed and predicted values of the dependent variable.

5. Interpretation of Coefficients: The coefficients obtained from the regression model provide insights into the relationship between the independent and dependent variables. A positive coefficient indicates a positive relationship (as the independent variable increases, the dependent variable also increases), while a negative coefficient indicates a negative relationship.

6. Model Evaluation: Once the model is fitted to the data, it is important to evaluate its performance. This can be done using metrics such as the coefficient of determination (R^2), which indicates the proportion of variance in the dependent variable explained by the independent variables.

4.3.6 Dataset with linear regression:

1. Dependent Variable (Y): In linear regression, the variable being predicted or explained is called the dependent variable. In the provided dataset, BFP is the dependent variable as it is the target variable being predicted.
2. Independent Variables (X): These are the variables used to predict the dependent variable. In the dataset, variables such as age, gender, height, weight, and BMI can serve as independent variables.

3. Linear Equation: The linear regression model seeks to find the equation of a straight line that best fits the relationship between the independent and dependent variables. The general form of a linear equation is:

$$[Y = \beta_0 + \beta_1 X_1 + \beta_2 X_2 + \dots + \beta_n X_n + \varepsilon]$$

4. Model Fitting: The goal of linear regression is to find the values of the coefficients ($\beta_0, \beta_1, \beta_2, \dots, \beta_n$) that minimize the sum of squared errors between the observed and predicted values of the dependent variable. This process is often performed using optimization techniques such as ordinary least squares (OLS) regression.

4.4 ALGORITHMS USED:

4.4.1 BODY COMPOSITION ANALYSIS:

- *Data Preprocessing:*

Dataset containing variables such as age, gender, height, weight, BMI, and body fat percentage (BFP). Processing Steps are to remove any missing or inconsistent data points. Normalize the numerical variables to ensure uniform scales across features. Encode categorical variables (e.g., gender) using one-hot encoding for numerical representation. To Capture images or video frames containing human subjects using a camera or webcam

- *Feature Selection:*

Identify relevant features that may influence body fat composition. Prioritize variables such as age, gender, height, weight, and BMI based on domain knowledge and statistical significance. Utilize a pre-trained face detection model, such as the Haar Cascade classifier, to detect faces within the captured images or frames. Extract bounding boxes or regions of interest (ROIs) corresponding to detected faces.

- *Model Training:*

Linear Regression: Choose linear regression as the predictive modeling technique due to its simplicity and interpretability. Separate the dataset into training and testing sets for model validation. Fit a linear regression model to the training data using ordinary least squares (OLS) regression. Assess the model's performance using evaluation metrics such as R^2 (coefficient of determination), root mean squared error (RMSE), and mean absolute error (MAE).

- *Model Evaluation:*

Performance Metrics is to Calculate R^2 , RMSE, and MAE to evaluate the goodness-of-fit and predictive accuracy of the linear regression model. Interpret the R^2 value to understand the proportion of variance in BFP explained by the independent variables. Analyze RMSE and MAE to assess the magnitude of errors between observed and predicted BFP values.

- *Model Interpretation:*

Examine the coefficients of the linear regression model to understand the magnitude and direction of the relationships between independent variables and BFP. Interpret the coefficients in the context of body fat composition analysis, considering factors such as age, gender, height, weight, and BMI. Utilize regression techniques or machine learning algorithms to establish a relationship between facial measurements and height.

- *Feature Extraction:*

Measure specific facial features relevant to height estimation, such as the distance between the chin and the top of the head or the distance between the eyes and the top

of the head. Compute additional features, if necessary, based on geometric relationships between facial landmarks.

- *Prediction and Deployment:*

Utilizing the trained linear regression model to predict body fat percentage for new individuals based on their demographic and anthropometric characteristics. Implement the model in a user-friendly interface for convenient access and utilization in healthcare, fitness, and wellness settings.

CHAPTER-5

PERFORMANCE ANALYSIS, RESULTS AND DISCUSSION

4.1 RESULTS AND DISCUSSION

5.1.1 HEIGHT AND WEIGHT DETECTION WITH ARDUINO:

The integration of the load cell and ultrasonic sensor with the Arduino Uno microcontroller has enabled the creation of an external parameters. In measurement system where the output is shown in Fig. 5.1 act as an External parameters which displayed on the LCD screen where the system calculates essential metrics such as weight, height, body mass index (BMI), and body fat percentage in real-time. This user-friendly interface provides individuals with instant access to their biometric data, facilitating informed decisions about their health and fitness.




FIG.5.1 HEIGHT AND WEIGHT DETECTION RESULT

5.1.2 BFP USING LINEAR REGRESSION:

In addition to utilizing the load cell and ultrasonic sensor for external parameter measurements, we also employed a sophisticated linear regression model which shown in Fig.5.2 to enhance the accuracy of body fat composition analysis. By using the collected datasets and demographic information, by training the model where it predicts the body fat percentage with remarkable precision. Based on the age and certain factors of individuals the BFP of a person is calculated. The implementation of the linear regression algorithm on the Arduino Uno microcontroller allows for real-time analysis of biometric data, Through the continuous refinement and validation, the linear regression model provides valuable insights into individuals health with high accuracy, offering personalized assessments and actionable recommendations.





FIG.5.2 BFP WITH LINEAR REGRESSION RESULT

5.1.3 BWP USING LINEAR REGRESSION:

In Fig.5.3 shows the linear regression method which determines the body water percentage (BWP) of individuals based on their height, weight, and body mass index (BMI). With the collected datasets and demographic information, the linear regression model accurately predicts the body water content, providing valuable insights into hydration levels and overall health. This approach offers a non-invasive and efficient means of assessing body water content, eliminating the need for invasive procedures or specialized equipment. Through continuous refinement and validation, the project aims to enhance the accuracy and reliability of the BWC prediction model, ultimately contributing to personalized health monitoring and wellness management.



```
def predict_bfp(age, height, weight, bmi):  
    # Reshape input data into a 2D array  
    input_data = [[age, height, weight, bmi]]  
  
    # Use the trained model to make predictions  
    predicted_bfp = model.predict(input_data)  
  
    return predicted_bfp[0]  
  
# Input values for AGE, HEIGHT, WEIGHT, and BMI  
age_input = float(input("Enter age: "))  
height_input = float(input("Enter height (in cm): "))  
weight_input = float(input("Enter weight (in kg): "))  
bmi_input = float(input("Enter BMI: "))  
  
# Predict BFP using the input values  
predicted_bfp = predict_bfp(age_input, height_input, weight_input, bmi_input)  
print("Predicted BFP:", predicted_bfp)  
print("predicted_bwp:", predicted_bwp)
```

Enter age: 19
Enter height (in cm): 152
Enter weight (in kg): 54
Enter BMI: 35.5263158
Predicted BFP: 29.873322972198583
predicted_bwp: 27.443684
/usr/local/lib/python3.10/dist-packages/sklearn/base.py:439: UserWarning: X does not have valid feature names, but LinearRegression was fitted with feature names
warnings.warn(

FIG.5.3 BWP WITH LINEAR REGRESSION RESULT

5.1.4. GRAPHICAL REPRESENTATION OF BFP:

Based on the persons external parameters the body fat of a person determined by using linear regression method in which the Fig.5.4 shows the graphical output the consistency of BFP level with the other three parameters such as (Height, Weight and BMI) graph. The graph could be generated based on a average of the persons BFP. The graph could be auto generated based on the model trained using the datasets to calculate the average body fat.




FIG.5.4. LINEAR REGRESSION GRAPHICAL REPRESENTATION

5.1.5 HEIGHT AND WEIGHT IN IMAGE PROCESSING:

In this method, the image processing algorithms is used to measure the height of a person accurately by using computer vision techniques. By employing a combination from facial landmark to toe of a person to detect and the system extracts key features from captured images to determine an individual's height. In Fig.5.5 shows the system calculates the distance between these points to estimate height with high accuracy. This innovative approach to height measurement eliminates the need for manual measurements and enables real-time assessment of biometric data. By using the image processing algorithms, the system offers a efficient solution for height measurements which leads to overcome the error in measurement of persons height.




FIG.5.5 HEIGHT DETECTION RESULT

5.1.6 WEIGHT DETECTION USING IMAGE PROCESSING:

We also used to calculate the persons body weight by using computer vision with the image processing which shown in Fig.5.6 done by utilizing facial features such as jawline length, eyebrow length, nose length, eye length, outer mouth length, and inner mouth length but the obtained results did not meet the desired level of accuracy. Factors such as variations in facial morphology, facial expressions, and camera angles may have introduced inconsistencies in the measurements, leading to discrepancies in the calculated weight. Additionally, limitations in the algorithms to handle diverse facial characteristics and environmental conditions may have impacted the accuracy of the results. the project underscores the importance of facial features for weight estimation and highlights the importance of further refinement and validation to improve the accuracy and reliability of the proposed approach. Future iterations of the algorithm may benefit from incorporating machine learning techniques to enhance its adaptability and effectiveness across a wider range of individuals and scenarios, ultimately leading to more precise and dependable weight estimation methods.




FIG.5.6 WEIGHT DETECTION RESULT

5.2. DISCUSSION:

The project aims to develop a comprehensive body composition measurement system that integrates external parameters like weight and height with internal parameters such as body water and fat content. It initiates with the creation of a non-real-time prototype utilizing ultrasonic sensors for height measurement and load cells for weight measurement. Through the derived formula, body mass index (BMI) is calculated from height and weight, facilitating the determination of fat percentage based on external parameters. Moreover, the project conducts linear regression analysis to establish predictive models using collected datasets, enhancing the precision of body fat composition assessment. The integration of technological advancements with established anthropometric measurements seeks to contribute to the development of more accurate and accessible methods for assessing body fat composition. Ethical considerations and privacy protections are prioritized throughout the project to ensure the responsible handling of sensitive health data. Ultimately, the goal is to create a versatile system with applications in healthcare, fitness assessment, and wellness monitoring, providing valuable insights into an individual's health and fitness status.

5.3. PERFORMANCE ANALYSIS:

The performance analysis of the body composition analysis system involves evaluating the accuracy, reliability, and efficiency of the implemented algorithms and sensor technologies for measuring height, weight, and body fat composition. Here's a breakdown of the performance analysis for each component: Height Detection: Evaluate the accuracy, precision, robustness, and speed of height detection algorithms using computer vision techniques. Weight Detection: Validate the accuracy, precision, sensitivity, and noise handling capabilities of the load cell for weight measurement. Body Fat Composition Analysis: Assess the prediction accuracy, model fit, generalization, and clinical relevance of linear regression models for predicting body fat percentage. Overall System Performance: Evaluate the integration, user experience, real-time monitoring capabilities, and scalability of the body composition analysis system for diverse applications in healthcare, fitness assessment, and

wellness monitoring. Ethical considerations and privacy protections are paramount throughout the performance analysis process to ensure responsible data handling and user privacy.

CHAPTER-6

SUMMARY AND CONCLUSION

6.1 SUMMARY:

This project aims to develop an advanced body composition measurement system that integrates cutting-edge sensor technologies and algorithmic techniques to accurately assess key biometric parameters such as height, weight, and body fat composition. By leveraging ultrasonic sensors for height measurement and load cells for weight measurement, the system provides precise and reliable data on external parameters, allowing for the calculation of body mass index (BMI) and subsequent estimation of fat percentage. Additionally, linear regression analysis is employed to establish predictive models based on collected datasets, enhancing the accuracy of body fat composition assessment. This integration of technological advancements with established anthropometric measurements contributes to the development of more precise and accessible methods for evaluating overall health and fitness.

Ethical considerations and privacy protections are prioritized throughout the project to ensure the responsible handling of sensitive health data. By adhering to ethical guidelines and implementing robust privacy measures, the project aims to safeguard user confidentiality and trust. Ultimately, the goal is to create a versatile system with applications in healthcare, fitness assessment, and wellness monitoring, providing valuable insights into an individual's health status and facilitating informed decision-making regarding lifestyle management and healthcare interventions.

The project also emphasizes the transition to real-time monitoring, incorporating body fat content analysis to assess body composition more accurately. This involves analyzing collected data to calculate body fat and water content, further enhancing the system's capabilities in providing comprehensive health assessments. Through these endeavors, the project seeks to empower individuals with actionable insights into their health and fitness, ultimately contributing to improved well-being and lifestyle management on a broader scale.

Furthermore, the project emphasizes the importance of continuous improvement and innovation in the field of body composition analysis. By conducting thorough performance analysis and validation, the system aims to identify areas for refinement and enhancement, ensuring its relevance and effectiveness in diverse real-world scenarios. Through ongoing research and development efforts, the project seeks to stay at the forefront of technological advancements in biometric measurement, ultimately delivering impactful solutions that empower individuals to make informed decisions about their health and well-being.

6.2 CONCLUSION:

In conclusion, the development of the comprehensive body composition measurement system represents a significant step forward in the field of biometric analysis. By integrating advanced sensor technologies with sophisticated algorithmic techniques, the project has laid the foundation for precise and accessible methods of assessing key biometric parameters such as height, weight, and body fat composition. Through meticulous calibration and validation processes, the system has demonstrated promising accuracy, reliability, and efficiency in providing valuable insights into an individual's health and fitness status.

Ethical considerations and privacy protections have been paramount throughout the project, underscoring the importance of responsible data handling and user confidentiality. By adhering to ethical guidelines and implementing robust privacy measures, the project aims to instill trust and confidence among users, ensuring the secure and ethical use of sensitive health data. This commitment to ethical standards further underscores the project's dedication to promoting user well-being and safeguarding their rights and privacy.

Moving forward, the project will continue to evolve and improve, leveraging ongoing research and development efforts to enhance the system's capabilities and usability. By incorporating user feedback and addressing any identified limitations or challenges,

the system aims to stay at the forefront of technological advancements in biometric analysis, delivering innovative solutions that meet the evolving needs of individuals and healthcare professionals alike. Ultimately, the project's ultimate goal is to empower individuals with actionable insights into their health and fitness, enabling them to make informed decisions and take proactive steps towards improving their overall well-being and lifestyle. Through collaboration, innovation, and a commitment to excellence, the project stands poised to make a meaningful impact in the field of biometric analysis and contribute to the advancement of personalized health management strategies.

REFERENCES

1. Antonios H. Tzamaloukas, Glen H. Murata, Dorothy J. Vanderjagt, Robert H. Glew; Estimates of body water, fat-free mass and body fat in patients on peritoneal dialysis by anthropometric formulas; *Kidney International*, Vol. 63 (2003), pp. 1605-1617.
2. Cyntia De Baptista, Ana Mattiello-Sverzut; Height and body mass index distribution in children and adolescents with and without spina bifida; July 2023 *Jornal de Pediatria* 100(1); DOI: 10.1016/j.jped.2023.07.002; License CC BY-NC-ND 4.0
3. Dongming Li, Linnan Huang, Yangrong Wen, Yueming Gao; Analysis of Electrical Impedance Myography Electrodes Configuration for Local Muscle Fatigue Evaluation Based on Finite Element Method; DOI: 10.1109/ACCESS.2020.3025150
4. Erika Aparecida Silveira, Larissa Silva Barbosa, Ana Paula Santos Rodrigues, Matias Noll, and Cesar De Oliveira. Body fat percentage assessment by skinfold equation, bioimpedance and densitometry in older adults; Silveira et al. *Archives of Public Health* (2020) 78:65.
5. Erika Severeyn, Sara Wong, Hector Herrera, Alexandra La Cruz, Jesus Velasquez, Monica Huerta; Prediction of Abnormal Body Fat Percentage by Anthropometric Parameters Using Receiver Operating Characteristic Curve. DOI: 978-1-7281-9365-6/20 (2020).
6. F. Javier Núñez, Diego Munguía-Izquierdo, and Luis Suárez-Arrones; Validity of Field Methods to Estimate Fat-Free Mass Changes Throughout the Season in Elite Youth Soccer Player; Published online 2020 Feb 12. doi: 10.3389/fphys.2020.00016

7. Francesco Campa, Stefania Toselli, Massimiliano Mazzilli, Luís Alberto Gobbo, Giuseppe Coratella; Assessment of Body Composition in Athletes: A Narrative Review of Available Methods with Special Reference to Quantitative and Qualitative Bioimpedance Analysis; PMID: 34065984 PMCID: PMC8150618 DOI: 10.3390/nu13051620.
8. Hanneke Pierre Fransicus Xavierus Moonen, Florianne Johannan Louise Van Zanyen, Luuk Driessen, Vivienne De Smet, Rianne Slingerland Boot, Marco Mensink, Arthur Raymond Hubert van Zanten; Association of bioelectrical impedance analysis body composition and disease severity in COVID 19 hospital ward and ICU patients; the Biac-19 study; Clinical Nutrition 40 (2021) 2328-2336; DOI: 10.1016/j.clnu.2020.10.023.
9. Jihan K Boughman, Melissa A Masters, Cass A Morgan, Tim M Ruden, Shannon G Rochelle; Assessing the Validity of Bioelectrical Impedance and Skinfold Calipers for Measuring Body Composition in NOLS Backcountry Hikers; PMID: 31495771 DOI: 10.1016/j.wem.2019.06.011
10. Jung, M.H., Namkoong, K., Lee, Y. et al. Wrist-wearable bioelectrical impedance analyzer with miniature electrodes for daily obesity management. Sci Rep 11, 1238 (2021). DOI: 10.1038/s41598-020-79667-3
11. Katherine Sullivan, Bjoern Hornikel, Clifton J. Holmes, Michael R. Esco & Michael V. Fedewa; Validity of a 3-compartment body composition model using body volume derived from a novel 2-dimensional image analysis program; DOI: <https://doi.org/10.1038/s41430-021-00899-1>
12. Lynn Moeng-Mahlangu, Makama A Monyeke, John J Reilly, Herculina S Kruger; Comparison of Several Prediction Equations Using Skinfold Thickness for Estimating Percentage Body Fat vs. Body Fat Percentage Determined by BIA in 6-8-Year-Old South African Children; PMID: 36361413 PMCID: PMC9656018 DOI: 10.3390/ijerph192114531.
13. Malia N M Blue, Grant M Tinsley, Eric D Ryan, Abbie E Smith-Ryan; Validity of Body-Composition Methods across Racial and Ethnic Populations; PMID: 33684215 PMCID: PMC8528114 DOI: 10.1093/advances/nmab016.

14. McPhee, P.G.; Singh, S.; Morrison, K.M. Childhood Obesity and Cardiovascular Disease Risk: Working Toward Solutions. *Can. J. Cardiol.* 2020, 36, 1352–1361.
15. Ningjian Wang, Ying Sun, Haojie Zhang, Chi Chen, Yuying Wang, Jihui Zhang, Fangzhen Xia, Christian Benedict, Xiao Tan, Yingli Lu; Total and regional fat-to-muscle mass ratio measured by bioelectrical impedance and risk of incident type 2 diabetes; Published online 30 September 2021 in Wiley Online Library. DOI: 10.1002/jcsm.12822.
16. Rodriguez-Martinez, A.; Zhou, B.; Sophiea, M.K.; Bentham, J.; Paciorek, C.J.; Iurilli, M.L.C.; Carrillo-Larco, M.R.; Bennett, J.E.; Di Cesare, M.; Taddei, C. Height and body-mass index trajectories of school-aged children and adolescents from 1985 to 2019 in 200 countries and territories: A pooled analysis of 2181 population-based studies with 65 million participants. *Lancet* 2020, 396, 1511–1524.
17. Sadowsky M, McConkey R, Shellard A.; Obesity in youth and adults with intellectual disability in Europe and Eurasia. *Journal of Applied Research in Intellectual Disabilities* Volume 33, Issue 2/ p. 321-326.
18. Seung-Chul Shin, Jinkyu Lee, Soyeon Choe, Hyuk In Yang, Jihee Min, Ki-Yong Ahn, Justin Y. Jeon and Hong-Goo Kang; Dry Electrode-Based Body Fat Estimation System with Anthropometric Data for Use in a Wearable Device; *Sensors* 2019, 19(9), 2177; <https://doi.org/10.3390/s19092177>
19. Sergey G. Rudnev and Elena Z. Godina, Studies on human body composition in Russia: past and present; *Journal of Physiological Anthropology* (2022) 41:18.
20. Shengkui Zhang, Han Wang, Yongbin Wang, Miao Yu, and Juxiang Yuan; Association of Rotating Night Shift Work with Body Fat Percentage and Fat Mass Index among.
21. Shengkui Zhang, Han Wang, Yongbin Wang, Miao Yu, and Juxiang Yuan; Association of Rotating Night Shift Work with Body Fat Percentage and Fat Mass Index among
22. Ugras, Seda (2020) Evaluating of altered hydration status on effectiveness of body composition analysis using bioelectric impedance analysis, *Libyan Journal of Medicine*, 15:1,1741904, DOI: 10.1080/19932820.2020.1741904.

23. Unhee Lim, Kristine R. Monroe, Steve Buchthal, Bo Fan, Iona Cheng, Bruce S. Kristal, Johanna W. Lamp, Meredith A. Hullar, Adrian A. Franke, Daniel O. Stram, Lynne R. Wilkens, John Shepherd, Thomas Ernst, and Loïc Le Marchand; Propensity for Intra-abdominal and Hepatic Adiposity Varies Among Ethnic Groups; *Gastroenterology*. 2019 March ; 156(4): 966–975.e10. doi:10.1053/j.gastro.2018.11.021.
24. Willem G.P.M. Looijaard, Sandra N. Stapel, Ingeborg M. Dekker, Hanna Rusticus, Sharon Remmelzwaal, Armand R.J. Girbes, Peter J.M. Weijs, Heleen M. Oudemans-van Straaten; Identifying critically ill patients with low muscle mass: Agreement between bioelectrical impedance analysis and computed tomography; *Clinical Nutrition* 39 (2020) 1809-1817.
25. Wang, Ningjian, Ying Sun, Haojie Zhang, Chi Chen, Yuying Wang, Jihui Zhang, Fangzhen Xia, Christian Benedict, Xiao Tan, and Yingli Lu. "Total and regional fat-to-muscle mass ratio measured by bioelectrical impedance and risk of incident type 2 diabetes." Published online 30 September 2021 in Wiley Online Library. DOI: 10.1002/jcsm.12822.
26. Wang, Shengkui, Han Wang, Yongbin Wang, Miao Yu, and Juxiang Yuan. "Association of Rotating Night Shift Work with Body Fat Percentage and Fat Mass Index among Female Steelworkers in North China." Published online 2021 Jun 11. doi: 10.3390/ijerph18126355.
27. Wang, Willem G.P.M. Looijaard, Sandra N. Stapel, Ingeborg M. Dekker, Hanna Rusticus, Sharon Remmelzwaal, Armand R.J. Girbes, Peter J.M. Weijs, Heleen M. Oudemans-van Straaten; Identifying critically ill patients with low muscle mass: Agreement between bioelectrical impedance analysis and computed tomography; *Clinical Nutrition* 39 (2020) 1809-1817.
28. Wang, Ugras, Seda (2020) Evaluating of altered hydration status on effectiveness of body composition analysis using bioelectric impedance analysis, *Libyan Journal of Medicine*, 15:1,1741904, DOI: 10.1080/19932820.2020.1741904.

29. Zhang, Shengkui, Han Wang, Yongbin Wang, Miao Yu, and Juxiang Yuan; Association of Rotating Night Shift Work with Body Fat Percentage and Fat Mass Index among Female Steelworkers in North China; Published online 2021 Jun 11. doi: 10.3390/ijerph18126355
30. Zhang, Sergey G. Rudnev and Elena Z. Godina, Studies on human body composition in Russia: past and present; Journal of Physiological Anthropology (2022) 41:18.